

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
 Data File : VN086629.D
 Acq On : 14 May 2025 16:09
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 15 03:05:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	199899	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	383298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	366514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	165907	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138233	47.688	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.380%
35) Dibromofluoromethane	8.159	113	112153	63.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	126.080%#
50) Toluene-d8	10.565	98	488472	51.378	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.760%
62) 4-Bromofluorobenzene	12.847	95	182811	52.717	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
Data File : VN086629.D
Acq On : 14 May 2025 16:09
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 15 03:05:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

