

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN051520\
 Data File : VN061411.D
 Acq On : 15 May 2020 15:06
 Operator : JC/MD
 Sample : VN0515WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0515WBL01

Quant Time: May 16 00:35:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N051420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 14 16:04:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	14753	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	85205	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	72652	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	39971	30.25	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.83%
60) 4-Bromofluorobenzene	12.38	95	32558	26.25	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	87.50%
63) Toluene-d8	10.06	98	106768	31.74	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN051520\
Data File : VN061411.D
Acq On : 15 May 2020 15:06
Operator : JC/MD
Sample : VN0515WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0515WBL01

Quant Time: May 16 00:35:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N051420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 14 16:04:10 2020
Response via : Initial Calibration

