

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051523\
 Data File : VN077680.D
 Acq On : 15 May 2023 20:46
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: May 16 02:50:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 02:47:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	321845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	591222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	523640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	204509	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	239308	49.694	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.380%
35) Dibromofluoromethane	8.171	113	189486	49.150	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	10.571	98	733163	49.407	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.853	95	237836	45.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051523\
Data File : VN077680.D
Acq On : 15 May 2023 20:46
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: May 16 02:50:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 02:47:47 2023
Response via : Initial Calibration

