

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051524\
 Data File : VN082137.D
 Acq On : 15 May 2024 17:38
 Operator : JC\MD
 Sample : VN0515WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0515WBL01

Quant Time: May 16 05:51:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104343	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	135817	53.024	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.040%
35) Dibromofluoromethane	8.165	113	91482	52.463	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.920%
50) Toluene-d8	10.571	98	358829	53.674	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.340%
62) 4-Bromofluorobenzene	12.853	95	117071	49.062	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	98.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051524\
Data File : VN082137.D
Acq On : 15 May 2024 17:38
Operator : JC\MD
Sample : VN0515WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0515WBL01

Quant Time: May 16 05:51:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

