

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051525\
 Data File : VN086647.D
 Acq On : 15 May 2025 16:00
 Operator : JC\MD
 Sample : PB167995ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167995ZHE#09

Quant Time: May 16 05:29:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	223563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	420929	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	410683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	166378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150898	46.547	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 93.100%			
35) Dibromofluoromethane	8.165	113	124254	63.603	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 127.200%#			
50) Toluene-d8	10.565	98	542015	51.913	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.820%			
62) 4-Bromofluorobenzene	12.847	95	192311	50.499	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.000%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	27976	19.703	ug/l	98
20) Methylene Chloride	5.277	84	7753	2.587	ug/l	96
43) Isopropyl Acetate	8.688	43	63566	6.882	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051525\
Data File : VN086647.D
Acq On : 15 May 2025 16:00
Operator : JC\MD
Sample : PB167995ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167995ZHE#09

Quant Time: May 16 05:29:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

