


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051622\
 Data File : VN072494.D
 Acq On : 16 May 2022 11:42
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N051622W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue May 17 06:34:50 2022



AutoFind: Averaged scan 1819 to 1835; Bkg corrected with scan 1818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	2010	PASS
75	95	30	60	56.9	5395	PASS
95	95	100	100	100.0	9479	PASS
96	95	5	9	5.6	533	PASS
173	174	0.00	2	1.0	74	PASS
174	95	50	100	81.9	7762	PASS
175	174	5	9	6.5	508	PASS
176	174	95	101	99.4	7712	PASS
177	176	5	9	6.5	505	PASS