

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
 Data File : VN077689.D
 Acq On : 16 May 2023 13:52
 Operator : JC\MD
 Sample : 02804-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-051523

Quant Time: May 17 04:00:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	374409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	684204	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	591486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	236736	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	267087	47.676	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.360%
35) Dibromofluoromethane	8.171	113	210556	47.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	10.570	98	836747	48.724	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.440%
62) 4-Bromofluorobenzene	12.853	95	280010	46.167	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
Data File : VN077689.D
Acq On : 16 May 2023 13:52
Operator : JC\MD
Sample : 02804-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-051523

Quant Time: May 17 04:00:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

