

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
 Data File : VN077690.D
 Acq On : 16 May 2023 14:15
 Operator : JC\MD
 Sample : 02803-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-02-051523

Quant Time: May 17 04:00:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	345991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	630050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	557051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	219087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	245879	47.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.000%
35) Dibromofluoromethane	8.177	113	204725	49.830	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.660%
50) Toluene-d8	10.571	98	786685	49.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.853	95	258674	46.315	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.620%
Target Compounds						
16) Acetone	4.453	43	5221	3.206	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
Data File : VN077690.D
Acq On : 16 May 2023 14:15
Operator : JC\MD
Sample : 02803-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-02-051523

Quant Time: May 17 04:00:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

