

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
 Data File : VN077701.D
 Acq On : 16 May 2023 18:38
 Operator : JC\MD
 Sample : 02798-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WATERWITCH-AVE-SAMPLE

Quant Time: May 17 04:08:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	347779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648562	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	592765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	227974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	249711	47.988	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.980%
35) Dibromofluoromethane	8.177	113	205489	48.588	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	10.577	98	812124	49.889	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.780%
62) 4-Bromofluorobenzene	12.859	95	267111	46.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.920%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	104731	63.972	ug/l	93
20) Methylene Chloride	5.295	84	25643	4.208	ug/l #	88
25) 2-Butanone	7.494	43	25558	9.980	ug/l #	80
37) Ethyl Acetate	7.571	43	70478	12.783	ug/l	96
43) Isopropyl Acetate	8.694	43	287231	31.894	ug/l #	84
54) cis-1,3-Dichloropropene	10.365	75	107788	14.480	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051623\
Data File : VN077701.D
Acq On : 16 May 2023 18:38
Operator : JC\MD
Sample : 02798-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WATERWITCH-AVE-SAMPLE

Quant Time: May 17 04:08:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

