

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082156.D
 Acq On : 16 May 2024 12:55
 Operator : JC\MD
 Sample : P2403-09 0.5PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024

Quant Time: May 17 04:10:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	308122	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	148177	54.112	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.220%	
35) Dibromofluoromethane	8.171	113	98965	52.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.420%	
50) Toluene-d8	10.571	98	360437	49.600	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.200%	
62) 4-Bromofluorobenzene	12.847	95	118045	45.512	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 91.020%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.130	85	874	0.350	ug/l	76
3) Chloromethane	2.359	50	1652	0.583	ug/l	96
4) Vinyl Chloride	2.512	62	1427	0.536	ug/l #	40
5) Bromomethane	2.959	94	1450	0.947	ug/l	82
6) Chloroethane	3.124	64	1103	0.634	ug/l #	85
7) Trichlorofluoromethane	3.500	101	1838	0.527	ug/l #	83
8) Diethyl Ether	3.971	74	338	0.272	ug/l #	11
9) 1,1,2-Trichlorotrifluo...	4.383	101	968	0.475	ug/l #	38
10) Methyl Iodide	4.600	142	679	0.331	ug/l #	40
11) Tert butyl alcohol	5.530	59	1029	2.341	ug/l #	71
12) 1,1-Dichloroethene	4.353	96	1107	0.584	ug/l #	61
13) Acrolein	4.183	56	512	1.572	ug/l	94
14) Allyl chloride	5.024	41	2064	0.509	ug/l #	85
15) Acrylonitrile	5.724	53	2664	2.340	ug/l	94
16) Acetone	4.436	43	3380	3.009	ug/l	93
17) Carbon Disulfide	4.730	76	3409	0.557	ug/l #	74
18) Methyl Acetate	5.030	43	1644	0.588	ug/l #	47
19) Methyl tert-butyl Ether	5.800	73	3418	0.494	ug/l #	90
20) Methylene Chloride	5.277	84	1856	0.819	ug/l #	75
21) trans-1,2-Dichloroethene	5.765	96	679	0.329	ug/l #	37
22) Diisopropyl ether	6.683	45	3603	0.426	ug/l #	79
23) Vinyl Acetate	6.612	43	16275	2.438	ug/l #	84
24) 1,1-Dichloroethane	6.577	63	2111	0.484	ug/l #	68
25) 2-Butanone	7.488	43	4429	2.601	ug/l	96
26) 2,2-Dichloropropane	7.483	77	1563	0.428	ug/l #	70
27) cis-1,2-Dichloroethene	7.488	96	1216	0.499	ug/l	76
28) Bromochloromethane	7.818	49	951	0.468	ug/l #	62
29) Tetrahydrofuran	7.847	42	2721	2.488	ug/l #	42
30) Chloroform	7.971	83	2220	0.533	ug/l	91
31) Cyclohexane	8.218	56	7349	1.680	ug/l #	37
32) 1,1,1-Trichloroethane	8.165	97	1824	0.493	ug/l #	53
36) 1,1-Dichloropropene	8.365	75	1658	0.522	ug/l #	57
37) Ethyl Acetate	7.565	43	1697	0.477	ug/l #	68
38) Carbon Tetrachloride	8.359	117	1602	0.497	ug/l #	85
39) Methylcyclohexane	9.600	83	1645	0.472	ug/l	91
40) Benzene	8.612	78	4663	0.492	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082156.D
 Acq On : 16 May 2024 12:55
 Operator : JC\MD
 Sample : P2403-09 0.5PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024

Quant Time: May 17 04:10:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	614	0.293	ug/l #	51
42) 1,2-Dichloroethane	8.665	62	2066	0.567	ug/l #	73
43) Isopropyl Acetate	8.694	43	3027	0.480	ug/l #	67
44) Trichloroethene	9.347	130	1018	0.473	ug/l	70
45) 1,2-Dichloropropane	9.624	63	1096	0.448	ug/l	90
46) Dibromomethane	9.706	93	823	0.513	ug/l	87
47) Bromodichloromethane	9.888	83	1487	0.439	ug/l #	92
48) Methyl methacrylate	9.688	41	1482	0.493	ug/l #	68
49) 1,4-Dioxane	9.694	88	365	8.672	ug/l #	46
51) 4-Methyl-2-Pentanone	10.453	43	7982	2.229	ug/l #	81
52) Toluene	10.629	92	2753	0.484	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	1437	0.410	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	1373	0.370	ug/l #	59
55) 1,1,2-Trichloroethane	11.012	97	1018	0.492	ug/l #	76
56) Ethyl methacrylate	10.871	69	1394	0.397	ug/l #	54
57) 1,3-Dichloropropane	11.165	76	1686	0.436	ug/l	90
58) 2-Chloroethyl Vinyl ether	10.159	63	3686	2.036	ug/l	91
59) 2-Hexanone	11.194	43	5075	1.935	ug/l	95
60) Dibromochloromethane	11.359	129	1071	0.469	ug/l	89
61) 1,2-Dibromoethane	11.459	107	824	0.388	ug/l	99
64) Tetrachloroethene	11.106	164	1101	0.549	ug/l	89
65) Chlorobenzene	11.888	112	2873	0.480	ug/l #	80
66) 1,1,1,2-Tetrachloroethane	11.965	131	943	0.479	ug/l #	57
67) Ethyl Benzene	11.965	91	4696	0.439	ug/l	93
68) m/p-Xylenes	12.076	106	3408	0.869	ug/l	96
69) o-Xylene	12.400	106	1328	0.350	ug/l	68
70) Styrene	12.412	104	2488	0.394	ug/l	88
71) Bromoform	12.570	173	627	0.443	ug/l #	85
73) Isopropylbenzene	12.694	105	3827	0.432	ug/l	94
74) N-amyl acetate	12.494	43	2126	0.436	ug/l #	78
75) 1,1,2,2-Tetrachloroethane	12.941	83	1132	0.443	ug/l #	91
76) 1,2,3-Trichloropropane	12.847	75	70321	29.947	ug/l #	1
77) Bromobenzene	12.982	156	1135	0.537	ug/l	91
78) n-propylbenzene	13.035	91	4950	0.472	ug/l	100
79) 2-Chlorotoluene	13.123	91	3179	0.482	ug/l	91
80) 1,3,5-Trimethylbenzene	13.170	105	3100	0.433	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.847	75	70321	70.044	ug/l #	12
82) 4-Chlorotoluene	13.223	91	3029	0.470	ug/l	89
83) tert-Butylbenzene	13.435	119	2926	0.467	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	3005	0.410	ug/l	94
85) sec-Butylbenzene	13.617	105	3571	0.411	ug/l	91
86) p-Isopropyltoluene	13.735	119	3274	0.464	ug/l	90
87) 1,3-Dichlorobenzene	13.741	146	1804	0.471	ug/l	92
88) 1,4-Dichlorobenzene	13.741	146	1804	0.462	ug/l	95
89) n-Butylbenzene	14.059	91	2592	0.408	ug/l	90
90) Hexachloroethane	14.335	117	479	0.366	ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	2012	0.555	ug/l	88
92) 1,2-Dibromo-3-Chloropr...	14.723	75	77	0.147	ug/l #	2
93) 1,2,4-Trichlorobenzene	15.394	180	990	0.488	ug/l	89
94) Hexachlorobutadiene	15.506	225	302	0.339	ug/l #	67
95) Naphthalene	15.647	128	2265	0.356	ug/l #	78
96) 1,2,3-Trichlorobenzene	15.841	180	718	0.353	ug/l	81

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082156.D
Acq On : 16 May 2024 12:55
Operator : JC\MD
Sample : P2403-09 0.5PPB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Quant Time: May 17 04:10:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion:168 Resp: 173419
Ion Ratio Lower Upper
168 100
99 63.8 48.2 72.4



#2
Dichlorodifluoromethane
Concen: 0.350 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 85 Resp: 874
Ion Ratio Lower Upper
85 100
87 44.7 15.8 47.4









#3
Chloromethane
Concen: 0.583 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 50 Resp: 1652
Ion Ratio Lower Upper
50 100
52 29.6 25.5 38.3



#4
Vinyl Chloride
Concen: 0.536 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 62 Resp: 1427
Ion Ratio Lower Upper
62 100
64 63.9 24.6 37.0#









#5
Bromomethane
Concen: 0.947 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 94 Resp: 1450
Ion Ratio Lower Upper
94 100
96 76.5 74.8 112.2



#6
Chloroethane
Concen: 0.634 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 64 Resp: 1103
Ion Ratio Lower Upper
64 100
66 39.1 24.7 37.1#









#7
Trichlorofluoromethane
Concen: 0.527 ug/l
RT: 3.500 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:101 Resp: 1838
Ion Ratio Lower Upper
101 100
103 49.7 50.6 75.8#



#8
Diethyl Ether
Concen: 0.272 ug/l
RT: 3.971 min Scan# 343
Delta R.T. 0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 74 Resp: 338
Ion Ratio Lower Upper
74 100
45 166.0 42.4 127.3#









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.475 ug/l
RT: 4.383 min Scan# 411
Delta R.T. 0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:101 Resp: 968
Ion Ratio Lower Upper
101 100
85 6.6 37.3 55.9#
151 21.2 59.2 88.8#



#10
Methyl Iodide
Concen: 0.331 ug/l
RT: 4.600 min Scan# 450
Delta R.T. 0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:142 Resp: 679
Ion Ratio Lower Upper
142 100
127 0.0 35.8 53.6#
141 0.0 11.4 17.2#





#11

Tert butyl alcohol

Concen: 2.341 ug/l

RT: 5.530 min Scan# 606

Delta R.T. 0.012 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion: 59 Resp: 1029

Ion Ratio Lower Upper

59 100

57 0.0 8.9 13.3#



#12

1,1-Dichloroethene

Concen: 0.584 ug/l

RT: 4.353 min Scan# 408

Delta R.T. 0.012 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 96 Resp: 1107


Ion Ratio Lower Upper

96 100

61 93.8 119.7 179.5#

98 41.3 49.6 74.4#





#13

Acrolein

Concen: 1.572 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion: 56 Resp: 512

Ion Ratio Lower Upper

56 100

55 65.0 55.8 83.8



#14

Allyl chloride

Concen: 0.509 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 41 Resp: 2064

Ion Ratio Lower Upper

41 100

39 80.5 65.4 98.0

76 20.1 37.3 55.9#









#15
Acrylonitrile
Concen: 2.340 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:	53	Resp:	2664
Ion	Ratio	Lower	Upper
53	100		
52	76.1	66.1	99.1
51	35.6	29.4	44.2



#16
Acetone
Concen: 3.009 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	43	Resp:	3380
Ion	Ratio	Lower	Upper
43	100		
58	27.4	25.2	37.8









#17
Carbon Disulfide
Concen: 0.557 ug/l
RT: 4.730 min Scan# 41
Delta R.T. 0.018 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 76 Resp: 3409
Ion Ratio Lower Upper
76 100
78 0.0 7.5 11.3#



#18
Methyl Acetate
Concen: 0.588 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 43 Resp: 1644
Ion Ratio Lower Upper
43 100
74 0.0 22.2 33.2#








#19
Methyl tert-butyl Ether
Concen: 0.494 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 73 Resp: 3418
Ion Ratio Lower Upper
73 100
57 25.1 16.5 24.7#



#20
Methylene Chloride
Concen: 0.819 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55


Tgt Ion: 84 Resp: 1856
Ion Ratio Lower Upper
84 100
49 76.4 84.9 127.3#
51 36.3 27.0 40.6
86 40.9 52.3 78.5#








#21
trans-1,2-Dichloroethene
Concen: 0.329 ug/l
RT: 5.765 min Scan# 64
Delta R.T. -0.024 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024







Tgt Ion: 96 Resp: 679
Ion Ratio Lower Upper
96 100
61 61.5 96.5 144.7#
98 0.0 51.3 76.9#



#22
Diisopropyl ether
Concen: 0.426 ug/l
RT: 6.683 min Scan# 804
Delta R.T. 0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 45 Resp: 3603
Ion Ratio Lower Upper
45 100
43 43.8 44.1 66.1#
87 18.6 27.0 40.6#
59 0.0 9.6 14.4#









#23
 Vinyl Acetate
 Concen: 2.438 ug/l
 RT: 6.612 min Scan# 791
 Delta R.T. 0.012 min
 Lab File: VN082156.D
 Acq: 16 May 2024 12:55




Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024


Tgt Ion: 43 Resp: 16275
 Ion Ratio Lower Upper
 43 100
 86 6.5 10.2 15.4#



#24
 1,1-Dichloroethane
 Concen: 0.484 ug/l
 RT: 6.577 min Scan# 786
 Delta R.T. 0.006 min
 Lab File: VN082156.D
 Acq: 16 May 2024 12:55

Tgt Ion: 63 Resp: 2111
 Ion Ratio Lower Upper
 63 100
 98 22.7 3.6 10.9#
 100 0.0 2.6 7.8#









#25
2-Butanone
Concen: 2.601 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Tgt Ion: 43 Resp: 4429
Ion Ratio Lower Upper
43 100
72 29.1 25.3 37.9



#26
2,2-Dichloropropane
Concen: 0.428 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 77 Resp: 1563
Ion Ratio Lower Upper
77 100
97 8.2 11.3 33.9#





#27

cis-1,2-Dichloroethene

Concen: 0.499 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN082156.D


Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024






Tgt Ion: 96 Resp: 1216

Ion Ratio Lower Upper

96 100

61 100.1 0.0 268.0

98 53.3 0.0 126.8



#28

Bromochloromethane

Concen: 0.468 ug/l

RT: 7.818 min Scan# 997

Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 49 Resp: 951

Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.6

130 48.1 66.7 100.1#





#29

Tetrahydrofuran

Concen: 2.488 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. 0.011 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion: 42 Resp: 2721

Ion Ratio Lower Upper

42 100

72 0.0 39.3 58.9#

71 17.2 37.4 56.0#



#30

Chloroform

Concen: 0.533 ug/l

RT: 7.971 min Scan# 1023

Delta R.T. 0.006 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55


Tgt Ion: 83 Resp: 2220

Ion Ratio Lower Upper

83 100

85 70.7 50.7 76.1





#31

Cyclohexane

Concen: 1.680 ug/l

RT: 8.218 min Scan# 1065

Delta R.T. -0.041 min

Lab File: VN082156.D


Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024






Tgt Ion: 56 Resp: 7349

Ion Ratio Lower Upper

56 100

69 143.2 24.4 36.6#

84 78.9 77.0 115.4



#32

1,1,1-Trichloroethane

Concen: 0.493 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 97 Resp: 1824

Ion Ratio Lower Upper

97 100

99 0.0 50.3 75.5#

61 45.9 36.6 55.0





#33

1,2-Dichloroethane-d4

Concen: 54.112 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion: 65 Resp: 148177

Ion Ratio Lower Upper

65 100

67 50.3 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 114 Resp: 308122


Ion Ratio Lower Upper

114 100

63 24.8 0.0 41.0

88 16.9 0.0 31.4









#35
Dibromofluoromethane
Concen: 52.213 ug/l
RT: 8.171 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:113 Resp: 98965
Ion Ratio Lower Upper
113 100
111 101.9 83.4 125.2
192 17.9 14.0 21.0



#36
1,1-Dichloropropene
Concen: 0.522 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 75 Resp: 1658
Ion Ratio Lower Upper
75 100
110 0.0 17.8 53.5#
77 18.8 25.0 37.4#





#37

Ethyl Acetate

Concen: 0.477 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN082156.D


Acq: 16 May 2024 12:55

Instrument :



MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024



Tgt Ion:	43	Resp:	1697
Ion	Ratio	Lower	Upper
43	100		
61	0.0	11.3	16.9#
70	0.0	8.7	13.1#



#38

Carbon Tetrachloride

Concen: 0.497 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Tgt Ion:	117	Resp:	1602
Ion	Ratio	Lower	Upper
117	100		
119	97.5	76.5	114.7
121	0.0	23.9	35.9#





#39

Methylcyclohexane

Concen: 0.472 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion: 83 Resp: 1645

Ion Ratio Lower Upper

83 100

55 71.9 53.0 79.6

98 39.0 37.1 55.7



#40

Benzene

Concen: 0.492 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. 0.006 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55

Tgt Ion: 78 Resp: 4663

Ion Ratio Lower Upper

78 100

77 25.8 20.0 30.0









#41
Methacrylonitrile
Concen: 0.293 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Tgt Ion:	41	Resp:	614
Ion	Ratio	Lower	Upper
41	100		
39	77.2	46.6	69.8#
67	21.2	67.3	100.9#
52	11.1	27.3	40.9#



#42
1,2-Dichloroethane
Concen: 0.567 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	62	Resp:	2066
Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	19.6









#43
Isopropyl Acetate
Concen: 0.480 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Tgt Ion: 43 Resp: 3027
Ion Ratio Lower Upper
43 100
61 6.6 22.8 34.2#
87 8.2 12.0 18.0#



#44
Trichloroethene
Concen: 0.473 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:130 Resp: 1018
Ion Ratio Lower Upper
130 100
95 68.4 0.0 195.4





#45

1,2-Dichloropropane

Concen: 0.448 ug/l

RT: 9.624 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion: 63 Resp: 1096

Ion Ratio Lower Upper

63 100

65 24.5 24.1 36.1



#46

Dibromomethane

Concen: 0.513 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion: 93 Resp: 823

Ion Ratio Lower Upper

93 100

95 70.8 64.8 97.2

174 77.3 73.4 110.0





#47

Bromodichloromethane

Concen: 0.439 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55





Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:	83	Resp:	1487
Ion	Ratio	Lower	Upper
83	100		
85	69.2	51.9	77.9
127	0.0	7.5	11.3



#48

Methyl methacrylate

Concen: 0.493 ug/l




RT: 9.688 min Scan# 1315


Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Tgt Ion:	41	Resp:	1482
Ion	Ratio	Lower	Upper
41	100		
69	46.8	77.9	116.9
39	61.1	50.9	76.3





#49

1,4-Dioxane

Concen: 8.672 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN082156.D


Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024






Tgt Ion: 88 Resp: 365

Ion Ratio Lower Upper

88 100

43 0.0 19.0 28.6#

58 19.2 50.6 76.0#



#50

Toluene-d8

Concen: 49.600 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55


Tgt Ion: 98 Resp: 360437

Ion Ratio Lower Upper

98 100

100 65.0 51.8 77.6









#51
4-Methyl-2-Pentanone
Concen: 2.229 ug/l
RT: 10.453 min Scan# 1445
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 43 Resp: 7982
Ion Ratio Lower Upper
43 100
58 29.5 33.2 49.8#



#52
Toluene
Concen: 0.484 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 92 Resp: 2753
Ion Ratio Lower Upper
92 100
91 173.9 136.7 205.1









#53
t-1,3-Dichloropropene
Concen: 0.410 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 75 Resp: 1437
Ion Ratio Lower Upper
75 100
77 33.6 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 0.370 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 75 Resp: 1373
Ion Ratio Lower Upper
75 100
77 37.4 25.0 37.4
39 83.8 34.8 52.2#









#55
1,1,2-Trichloroethane
Concen: 0.492 ug/l
RT: 11.012 min Scan# 1541
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024


Tgt Ion:	97	Resp:	1018
Ion	Ratio	Lower	Upper
97	100		
83	50.0	70.4	105.6#
85	47.2	45.6	68.4
99	71.4	50.9	76.3



#56
Ethyl methacrylate
Concen: 0.397 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	69	Resp:	1394
Ion	Ratio	Lower	Upper
69	100		
41	87.5	46.6	70.0#
39	65.5	26.5	39.7#





#57

1,3-Dichloropropane

Concen: 0.436 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion: 76 Resp: 1686

Ion Ratio Lower Upper

76 100

78 26.8 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 2.036 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55


Tgt Ion: 63 Resp: 3686

Ion Ratio Lower Upper

63 100

106 26.0 24.6 36.8









#59
2-Hexanone
Concen: 1.935 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 43 Resp: 5075
Ion Ratio Lower Upper
43 100
58 53.1 28.3 84.9



#60
Dibromochloromethane
Concen: 0.469 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 129 Resp: 1071
Ion Ratio Lower Upper
129 100
127 66.9 38.3 114.8









#61
1,2-Dibromoethane
Concen: 0.388 ug/l
RT: 11.459 min Scan# 1616
Delta R.T. -0.012 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:107 Resp: 824
Ion Ratio Lower Upper
107 100
109 95.4 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 45.512 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 95 Resp: 118045
Ion Ratio Lower Upper
95 100
174 76.1 0.0 160.6
176 73.2 0.0 152.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:117 Resp: 270797
Ion Ratio Lower Upper
117 100
82 61.9 44.6 66.8
119 32.1 25.9 38.9



#64
Tetrachloroethene
Concen: 0.549 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:164 Resp: 1101
Ion Ratio Lower Upper
164 100
166 139.7 97.6 146.4
129 82.2 73.7 110.5
131 88.2 75.8 113.6









#65
Chlorobenzene
Concen: 0.480 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Tgt Ion:112 Resp: 2873
Ion Ratio Lower Upper
112 100
114 43.2 25.7 38.5#



#66
1,1,1,2-Tetrachloroethane
Concen: 0.479 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:131 Resp: 943
Ion Ratio Lower Upper
131 100
133 79.3 47.5 142.6
119 0.0 32.9 98.7#









#67
Ethyl Benzene
Concen: 0.439 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 91 Resp: 4696
Ion Ratio Lower Upper
91 100
106 27.8 25.3 37.9



#68
m/p-Xylenes
Concen: 0.869 ug/l
RT: 12.076 min Scan# 1721
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:106 Resp: 3408
Ion Ratio Lower Upper
106 100
91 205.4 159.4 239.0









#69
o-Xylene
Concen: 0.350 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:106 Resp: 1328
Ion Ratio Lower Upper
106 100
91 259.9 104.7 314.1



#70
Styrene
Concen: 0.394 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:104 Resp: 2488
Ion Ratio Lower Upper
104 100
78 55.3 38.6 58.0
103 64.4 43.8 65.6





#71

Bromoform

Concen: 0.443 ug/l

RT: 12.570 min Scan# 1805

Delta R.T. -0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion:173 Resp: 627

Ion Ratio Lower Upper

173 100

175 38.8 24.6 74.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion:152 Resp: 111566


Ion Ratio Lower Upper

152 100

115 61.1 30.4 91.3

150 157.2 0.0 353.2





#73

Isopropylbenzene

Concen: 0.432 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :





MDL-WATER-03-QT2-2024

Tgt Ion:105 Resp: 3827

Ion Ratio Lower Upper

105 100

120 23.7 13.4 40.1



#74

N-ethyl acetate

Concen: 0.436 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Tgt Ion: 43 Resp: 2126




Ion Ratio Lower Upper


43 100

70 32.6 39.0 58.4#

55 18.5 21.3 31.9#

61 13.5 20.5 30.7#





#75

1,1,2,2-Tetrachloroethane

Concen: 0.443 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion: 83 Resp: 1132

Ion Ratio Lower Upper

83 100

131 0.0 5.5 16.4#

85 60.8 32.9 98.6



#76

1,2,3-Trichloropropane

Concen: 29.947 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.147 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55


Tgt Ion: 75 Resp: 70321

Ion Ratio Lower Upper

75 100

77 1.3 90.5 271.3#





#77

Bromobenzene

Concen: 0.537 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55





Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:	156	Resp:	1135
Ion	Ratio	Lower	Upper
156	100		
77	210.1	100.2	300.6
158	84.1	49.1	147.3



#78

n-propylbenzene

Concen: 0.472 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. -0.000 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Tgt Ion:	91	Resp:	4950
Ion	Ratio	Lower	Upper
91	100		
120	23.2	11.7	35.0









#79
2-Chlorotoluene
Concen: 0.482 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion: 91 Resp: 3179
Ion Ratio Lower Upper
91 100
126 29.9 17.5 52.5



#80
1,3,5-Trimethylbenzene
Concen: 0.433 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion: 105 Resp: 3100
Ion Ratio Lower Upper
105 100
120 48.4 24.8 74.3





#81

trans-1,4-Dichloro-2-butene

Concen: 70.044 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.112 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion: 75 Resp: 70321

Ion Ratio Lower Upper

75 100

53 0.0 72.9 109.3#

89 0.0 39.1 58.7#



#82

4-Chlorotoluene

Concen: 0.470 ug/l

RT: 13.223 min Scan# 1916

Delta R.T. -0.000 min

Lab File: VN082156.D




Acq: 16 May 2024 12:55


Tgt Ion: 91 Resp: 3029

Ion Ratio Lower Upper

91 100

126 28.8 17.4 52.3









#83
tert-Butylbenzene
Concen: 0.467 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:	119	Resp:	2926
Ion	Ratio	Lower	Upper
119	100		
91	67.5	32.6	97.8
134	22.9	12.4	37.0



#84
1,2,4-Trimethylbenzene
Concen: 0.410 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	105	Resp:	3005
Ion	Ratio	Lower	Upper
105	100		
120	49.7	22.9	68.7









#85
 sec-Butylbenzene
 Concen: 0.411 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN082156.D
 Acq: 16 May 2024 12:55




Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024


Tgt Ion:105 Resp: 3571
 Ion Ratio Lower Upper
 105 100
 134 16.3 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 0.464 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.006 min
 Lab File: VN082156.D
 Acq: 16 May 2024 12:55

Tgt Ion:119 Resp: 3274
 Ion Ratio Lower Upper
 119 100
 134 23.5 13.2 39.5
 91 16.2 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 0.471 ug/l

RT: 13.741 min Scan# 2004

Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024





Tgt Ion:146 Resp: 1804

Ion Ratio Lower Upper

146 100

111 37.0 20.8 62.5

148 57.3 32.1 96.3



#88

1,4-Dichlorobenzene

Concen: 0.462 ug/l

RT: 13.741 min Scan# 2004

Delta R.T. -0.071 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion:146 Resp: 1804


Ion Ratio Lower Upper

146 100

111 37.0 19.3 57.9

148 57.3 31.3 93.9









#89
n-Butylbenzene
Concen: 0.408 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024


Tgt Ion:	91	Resp:	2592
Ion	Ratio	Lower	Upper
91	100		
92	50.2	26.2	78.5
134	13.5	13.2	39.5



#90
Hexachloroethane
Concen: 0.366 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	117	Resp:	479
Ion	Ratio	Lower	Upper
117	100		
201	63.3	33.3	99.8









#91
1,2-Dichlorobenzene
Concen: 0.555 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55




Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024


Tgt Ion:	146	Resp:	2012
Ion	Ratio	Lower	Upper
146	100		
111	34.4	21.3	64.0
148	55.0	31.8	95.4



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.147 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:	75	Resp:	77
Ion	Ratio	Lower	Upper
75	100		
155	0.0	42.6	128.0#
157	0.0	53.0	159.0#





#93

1,2,4-Trichlorobenzene

Concen: 0.488 ug/l

RT: 15.394 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN082156.D


Acq: 16 May 2024 12:55

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024






Tgt Ion:180 Resp: 990

Ion Ratio Lower Upper

180 100

182 79.9 46.7 140.0

145 33.5 16.2 48.6



#94

Hexachlorobutadiene

Concen: 0.339 ug/l

RT: 15.506 min Scan# 2304

Delta R.T. 0.006 min

Lab File: VN082156.D

Acq: 16 May 2024 12:55




Tgt Ion:225 Resp: 302


Ion Ratio Lower Upper

225 100

223 23.2 31.1 93.2#

227 50.0 30.6 91.6









#95
Naphthalene
Concen: 0.356 ug/l
RT: 15.647 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Tgt Ion:128 Resp: 2265
Ion Ratio Lower Upper
128 100
127 6.2 10.1 15.1#
129 0.0 8.9 13.3#



#96
1,2,3-Trichlorobenzene
Concen: 0.353 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN082156.D
Acq: 16 May 2024 12:55

Tgt Ion:180 Resp: 718
Ion Ratio Lower Upper
180 100
182 113.2 46.0 137.8
145 40.8 17.1 51.3

