

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082154.D
 Acq On : 16 May 2024 12:08
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 17 05:01:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	287031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133553	51.502	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.000%	
35) Dibromofluoromethane	8.165	113	92330	52.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.580%	
50) Toluene-d8	10.571	98	363101	53.638	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.280%	
62) 4-Bromofluorobenzene	12.853	95	118954	49.232	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 98.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082154.D
Acq On : 16 May 2024 12:08
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 17 05:01:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

