

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082161.D
 Acq On : 16 May 2024 14:55
 Operator : JC\MD
 Sample : P2502-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240515057-01-TRIP-BLANK

Quant Time: May 17 05:02:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301438	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	277545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	141682	51.389	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.780%
35) Dibromofluoromethane	8.171	113	97536	52.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.200%
50) Toluene-d8	10.570	98	388496	54.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.300%
62) 4-Bromofluorobenzene	12.847	95	127006	50.052	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082161.D
Acq On : 16 May 2024 14:55
Operator : JC\MD
Sample : P2502-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240515057-01-TRIP-BLANK

Quant Time: May 17 05:02:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

