

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082165.D
 Acq On : 16 May 2024 16:30
 Operator : JC\MD
 Sample : PB160937TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160937TB

Quant Time: May 17 05:04:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	305900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	290044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138684	51.078	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.160%	
35) Dibromofluoromethane	8.171	113	98010	52.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.160%	
50) Toluene-d8	10.565	98	394497	54.681	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.360%	
62) 4-Bromofluorobenzene	12.847	95	132543	51.472	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 102.940%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	37455	33.628	ug/l	93
20) Methylene Chloride	5.271	84	15864	7.059	ug/l	88
37) Ethyl Acetate	7.565	43	5600	1.586	ug/l #	83
43) Isopropyl Acetate	8.688	43	607230	97.027	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082165.D
Acq On : 16 May 2024 16:30
Operator : JC\MD
Sample : PB160937TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160937TB

Quant Time: May 17 05:04:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

