

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082166.D
 Acq On : 16 May 2024 16:54
 Operator : JC\MD
 Sample : PB160937ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160937ZHE#01

Quant Time: May 17 05:04:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	177768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	325053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	309256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	149202	53.153	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.300%	
35) Dibromofluoromethane	8.171	113	103175	51.598	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.200%	
50) Toluene-d8	10.571	98	417652	54.480	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	108.960%	
62) 4-Bromofluorobenzene	12.853	95	139852	51.111	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	102.220%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	26555	23.062	ug/l	91
20) Methylene Chloride	5.283	84	10458	4.501	ug/l #	81
37) Ethyl Acetate	7.565	43	4785	1.275	ug/l #	82
43) Isopropyl Acetate	8.688	43	952908	143.290	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082166.D
Acq On : 16 May 2024 16:54
Operator : JC\MD
Sample : PB160937ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160937ZHE#01

Quant Time: May 17 05:04:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

