

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082170.D
 Acq On : 16 May 2024 18:30
 Operator : JC\MD
 Sample : PB160937ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160937ZHE#05

Quant Time: May 17 05:06:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	172220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	140339	51.607	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.220%		
35) Dibromofluoromethane	8.171	113	98427	52.961	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.920%		
50) Toluene-d8	10.571	98	388520	54.527	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.060%		
62) 4-Bromofluorobenzene	12.853	95	127662	50.198	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 100.400%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	26857	24.075	ug/l	94
20) Methylene Chloride	5.277	84	11142	4.950	ug/l #	81
43) Isopropyl Acetate	8.688	43	619921	100.295	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082170.D
Acq On : 16 May 2024 18:30
Operator : JC\MD
Sample : PB160937ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160937ZHE#05

Quant Time: May 17 05:06:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

