

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082172.D
 Acq On : 16 May 2024 19:18
 Operator : JC\MD
 Sample : PB160937ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160937ZHE#07

Quant Time: May 17 05:07:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	315115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	141607	51.566	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.140%	
35) Dibromofluoromethane	8.165	113	99670	51.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.840%	
50) Toluene-d8	10.571	98	404457	54.422	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.840%	
62) 4-Bromofluorobenzene	12.847	95	135020	50.901	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.800%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	20696	18.372	ug/l	95
20) Methylene Chloride	5.277	84	9138	4.020	ug/l #	73
43) Isopropyl Acetate	8.688	43	489681	75.956	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082172.D
Acq On : 16 May 2024 19:18
Operator : JC\MD
Sample : PB160937ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160937ZHE#07

Quant Time: May 17 05:07:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

