

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082173.D
 Acq On : 16 May 2024 19:42
 Operator : JC\MD
 Sample : P2497-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 280856

Quant Time: May 17 05:07:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303330	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142144	53.496	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.000%		
35) Dibromofluoromethane	8.171	113	94824	50.818	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.640%		
50) Toluene-d8	10.565	98	389297	54.417	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.840%		
62) 4-Bromofluorobenzene	12.847	95	125582	49.182	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 98.360%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	17985	16.500	ug/l	93
18) Methyl Acetate	5.030	43	3748	1.381	ug/l #	60
20) Methylene Chloride	5.283	84	9649	4.388	ug/l #	70
43) Isopropyl Acetate	8.688	43	467616	75.351	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082173.D
Acq On : 16 May 2024 19:42
Operator : JC\MD
Sample : P2497-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
280856

Quant Time: May 17 05:07:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

