

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082175.D
 Acq On : 16 May 2024 20:30
 Operator : JC\MD
 Sample : P2497-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT3997

Quant Time: May 17 05:08:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	183265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	326442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150002	51.836	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.680%		
35) Dibromofluoromethane	8.165	113	103271	51.427	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.860%		
50) Toluene-d8	10.571	98	416637	54.116	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.240%		
62) 4-Bromofluorobenzene	12.847	95	138143	50.271	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 100.540%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	17313	14.584	ug/l #	85
20) Methylene Chloride	5.289	84	10598	4.425	ug/l #	82
43) Isopropyl Acetate	8.688	43	419766	62.852	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082175.D
Acq On : 16 May 2024 20:30
Operator : JC\MD
Sample : P2497-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT3997

Quant Time: May 17 05:08:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

