

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082157.D
 Acq On : 16 May 2024 13:19
 Operator : JC\MD
 Sample : P2403-09 0.75PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024

Quant Time: May 17 04:11:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh
 Dadoda

05/17/2024
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107860	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	143969	52.364	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.720%	
35) Dibromofluoromethane	8.165	113	94456	50.351	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.700%	
50) Toluene-d8	10.565	98	350270	48.701	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.400%	
62) 4-Bromofluorobenzene	12.853	95	112231	43.719	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 87.440%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	1609	0.641	ug/l	92
3) Chloromethane	2.365	50	2453	0.862	ug/l	96
4) Vinyl Chloride	2.512	62	1657	0.620	ug/l	96
5) Bromomethane	2.965	94	1765	1.148	ug/l	90
6) Chloroethane	3.124	64	1472	0.843	ug/l	89
7) Trichlorofluoromethane	3.500	101	2548	0.728	ug/l	99
8) Diethyl Ether	3.971	74	772	0.619	ug/l #	50
9) 1,1,2-Trichlorotrifluo...	4.371	101	1534	0.750	ug/l #	83
10) Methyl Iodide	4.583	142	1151	0.560	ug/l #	84
11) Tert butyl alcohol	5.524	59	1689m	3.827	ug/l	
12) 1,1-Dichloroethene	4.341	96	1419	0.746	ug/l	93
13) Acrolein	4.183	56	955	2.920	ug/l #	68
14) Allyl chloride	5.024	41	2819	0.692	ug/l #	84
15) Acrylonitrile	5.718	53	4074	3.564	ug/l #	86
16) Acetone	4.436	43	4219	3.741	ug/l #	78
17) Carbon Disulfide	4.706	76	4669	0.760	ug/l #	74
18) Methyl Acetate	5.041	43	2152	0.766	ug/l #	52
19) Methyl tert-butyl Ether	5.794	73	5157	0.742	ug/l	92
20) Methylene Chloride	5.288	84	2302m	1.012	ug/l	
21) trans-1,2-Dichloroethene	5.783	96	1339	0.646	ug/l	98
22) Diisopropyl ether	6.677	45	5448	0.641	ug/l #	81
23) Vinyl Acetate	6.600	43	20551	3.066	ug/l #	90
24) 1,1-Dichloroethane	6.571	63	3176	0.725	ug/l #	81
25) 2-Butanone	7.482	43	5735	3.354	ug/l	95
26) 2,2-Dichloropropane	7.482	77	2481	0.676	ug/l #	74
27) cis-1,2-Dichloroethene	7.494	96	1768	0.723	ug/l	78
28) Bromochloromethane	7.806	49	1532	0.750	ug/l #	62
29) Tetrahydrofuran	7.847	42	3744	3.410	ug/l #	75
30) Chloroform	7.965	83	3087	0.738	ug/l	98
31) Cyclohexane	8.229	56	7585	1.727	ug/l #	35
32) 1,1,1-Trichloroethane	8.171	97	2765	0.745	ug/l #	51
36) 1,1-Dichloropropene	8.382	75	2046	0.650	ug/l #	41
37) Ethyl Acetate	7.565	43	2514	0.714	ug/l #	71
38) Carbon Tetrachloride	8.365	117	2267	0.711	ug/l #	73
39) Methylcyclohexane	9.600	83	1879	0.545	ug/l #	75
40) Benzene	8.606	78	6242	0.665	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
 Data File : VN082157.D
 Acq On : 16 May 2024 13:19
 Operator : JC\MD
 Sample : P2403-09 0.75PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024

Quant Time: May 17 04:11:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 04:08:43 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh
 Dadoda

05/17/2024
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	1541m	0.744	ug/l	
42) 1,2-Dichloroethane	8.677	62	2523	0.699	ug/l #	73
43) Isopropyl Acetate	8.688	43	3849	0.617	ug/l #	60
44) Trichloroethene	9.347	130	1629	0.764	ug/l	87
45) 1,2-Dichloropropane	9.629	63	1687	0.697	ug/l	90
46) Dibromomethane	9.712	93	1186	0.746	ug/l #	78
47) Bromodichloromethane	9.888	83	2147	0.640	ug/l #	96
48) Methyl methacrylate	9.676	41	2019	0.679	ug/l #	75
49) 1,4-Dioxane	9.700	88	582	13.971	ug/l #	64
51) 4-Methyl-2-Pentanone	10.447	43	11450	3.230	ug/l #	86
52) Toluene	10.629	92	3932	0.698	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	2286	0.659	ug/l #	83
54) cis-1,3-Dichloropropene	10.312	75	2536	0.690	ug/l #	83
55) 1,1,2-Trichloroethane	11.018	97	1481	0.724	ug/l #	81
56) Ethyl methacrylate	10.876	69	1951	0.562	ug/l #	54
57) 1,3-Dichloropropane	11.170	76	2735	0.715	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	5637	3.146	ug/l	91
59) 2-Hexanone	11.200	43	8115	3.126	ug/l	87
60) Dibromochloromethane	11.359	129	1449	0.642	ug/l	91
61) 1,2-Dibromoethane	11.465	107	1418	0.675	ug/l #	77
64) Tetrachloroethene	11.106	164	1507	0.767	ug/l #	89
65) Chlorobenzene	11.894	112	4203	0.718	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	1116	0.579	ug/l #	57
67) Ethyl Benzene	11.965	91	6407	0.612	ug/l	97
68) m/p-Xylenes	12.070	106	4809	1.253	ug/l	91
69) o-Xylene	12.400	106	2403	0.646	ug/l	94
70) Styrene	12.412	104	3312	0.536	ug/l	96
71) Bromoform	12.576	173	795	0.574	ug/l #	99
73) Isopropylbenzene	12.694	105	6080	0.710	ug/l	98
74) N-amyl acetate	12.500	43	3059	0.649	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.941	83	2186	0.885	ug/l #	77
76) 1,2,3-Trichloropropane	12.994	75	1937m	0.853	ug/l	
77) Bromobenzene	12.982	156	1833	0.898	ug/l	84
78) n-propylbenzene	13.035	91	6783	0.668	ug/l	89
79) 2-Chlorotoluene	13.129	91	4463	0.700	ug/l	95
80) 1,3,5-Trimethylbenzene	13.176	105	4290	0.619	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.741	75	442m	0.455	ug/l	
82) 4-Chlorotoluene	13.223	91	4281	0.687	ug/l	87
83) tert-Butylbenzene	13.441	119	4060	0.670	ug/l	93
84) 1,2,4-Trimethylbenzene	13.482	105	4310	0.608	ug/l	91
85) sec-Butylbenzene	13.617	105	5443	0.648	ug/l	99
86) p-Isopropyltoluene	13.729	119	3832	0.561	ug/l #	88
87) 1,3-Dichlorobenzene	13.729	146	2571	0.695	ug/l	92
88) 1,4-Dichlorobenzene	13.811	146	2714m	0.719	ug/l	
89) n-Butylbenzene	14.053	91	3663	0.597	ug/l	92
90) Hexachloroethane	14.335	117	841	0.665	ug/l	78
91) 1,2-Dichlorobenzene	14.111	146	2423	0.692	ug/l	90
92) 1,2-Dibromo-3-Chloropr...	14.723	75	172	0.339	ug/l #	55
93) 1,2,4-Trichlorobenzene	15.388	180	1227	0.625	ug/l	92
94) Hexachlorobutadiene	15.500	225	616	0.715	ug/l	79
95) Naphthalene	15.641	128	3336	0.542	ug/l #	84
96) 1,2,3-Trichlorobenzene	15.847	180	1152	0.586	ug/l	82

05/17/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051624\
Data File : VN082157.D
Acq On : 16 May 2024 13:19
Operator : JC\MD
Sample : P2403-09 0.75PPB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Quant Time: May 17 04:11:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 04:08:43 2024
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024

Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda

05/17/2024
Supervised By :Semsettin
Yesilyurt

05/17/2024


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion:168 Resp: 174120
Ion Ratio Lower Upper
168 100
99 63.3 48.2 72.4

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt

05/17/2024



#2
Dichlorodifluoromethane
Concen: 0.641 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 85 Resp: 1609
Ion Ratio Lower Upper
85 100
87 27.4 15.8 47.4





#3
Chloromethane
Concen: 0.862 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 50 Resp: 245
Ion Ratio Lower Upper
50 100
52 33.9 25.5 38.3

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#4
Vinyl Chloride
Concen: 0.620 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 62 Resp: 1657
Ion Ratio Lower Upper
62 100
64 32.9 24.6 37.0





#5
Bromomethane
Concen: 1.148 ug/l
RT: 2.965 min Scan# 176
Delta R.T. 0.012 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 94 Resp: 176
Ion Ratio Lower Upper
94 100
96 84.2 74.8 112.2

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#6
Chloroethane
Concen: 0.843 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 64 Resp: 1472
Ion Ratio Lower Upper
64 100
66 36.9 24.7 37.1





#7
 Trichlorofluoromethane
 Concen: 0.728 ug/l
 RT: 3.500 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024





Tgt Ion:101 Resp: 2548
 Ion Ratio Lower Upper
 101 100
 103 63.6 50.6 75.8

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/17/2024
 Supervised By :Semsettin
 Yesilyurt


05/17/2024



#8
 Diethyl Ether
 Concen: 0.619 ug/l
 RT: 3.971 min Scan# 343
 Delta R.T. 0.012 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19


Tgt Ion: 74 Resp: 772
 Ion Ratio Lower Upper
 74 100
 45 130.8 42.4 127.3#





#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.750 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion:101 Resp: 1534
Ion Ratio Lower Upper
101 100
85 41.9 37.3 55.9
151 54.6 59.2 88.8

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#10
Methyl Iodide
Concen: 0.560 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:142 Resp: 1151
Ion Ratio Lower Upper
142 100
127 38.5 35.8 53.6
141 0.0 11.4 17.2





#11
Tert butyl alcohol
Concen: 3.827 ug/l m
RT: 5.524 min Scan# 607
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 59 Resp: 1689
Ion Ratio Lower Upper
59 100
57 3.3 8.9 13.3

Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda

05/17/2024
Supervised By :Semsettin
Yesilyurt



05/17/2024




#12
1,1-Dichloroethene
Concen: 0.746 ug/l
RT: 4.341 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19




Tgt Ion: 96 Resp: 1419
Ion Ratio Lower Upper
96 100
61 143.9 119.7 179.5
98 72.9 49.6 74.4





#13
Acrolein
Concen: 2.920 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 56 Resp: 95
Ion Ratio Lower Upper
56 100
55 43.8 55.8 83.8

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt

05/17/2024



#14
Allyl chloride
Concen: 0.692 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 41 Resp: 2819
Ion Ratio Lower Upper
41 100
39 83.7 65.4 98.0
76 20.3 37.3 55.9





#15
Acrylonitrile
Concen: 3.564 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 53 Resp: 4074
Ion Ratio Lower Upper
53 100
52 80.2 66.1 99.1
51 13.8 29.4 44.2

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#16
Acetone
Concen: 3.741 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.012 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 43 Resp: 4219
Ion Ratio Lower Upper
43 100
58 43.6 25.2 37.8





#17
Carbon Disulfide
Concen: 0.760 ug/l
RT: 4.706 min Scan# 469
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion: 76 Resp: 4669
Ion Ratio Lower Upper
76 100
78 0.0 7.5 11.3#

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#18
Methyl Acetate
Concen: 0.766 ug/l
RT: 5.041 min Scan# 525
Delta R.T. 0.017 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 43 Resp: 2152
Ion Ratio Lower Upper
43 100
74 2.6 22.2 33.2#





#19
Methyl tert-butyl Ether
Concen: 0.742 ug/l
RT: 5.794 min Scan# 615
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 73 Resp: 515
Ion Ratio Lower Upper
73 100
57 24.1 16.5 24.7

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#20
Methylene Chloride
Concen: 1.012 ug/l m
RT: 5.288 min Scan# 567
Delta R.T. 0.012 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 84 Resp: 2302
Ion Ratio Lower Upper
84 100
49 103.3 84.9 127.3
51 43.2 27.0 40.6#
86 56.8 52.3 78.5





#21
trans-1,2-Dichloroethene
Concen: 0.646 ug/l
RT: 5.783 min Scan# 613
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion: 96 Resp: 1339
Ion Ratio Lower Upper
96 100
61 117.9 96.5 144.7
98 64.7 51.3 76.9

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#22
Diisopropyl ether
Concen: 0.641 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 45 Resp: 5448
Ion Ratio Lower Upper
45 100
43 45.7 44.1 66.1
87 17.5 27.0 40.6#
59 18.2 9.6 14.4#





#23
 Vinyl Acetate
 Concen: 3.066 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19

Instrument : MSVOA_N
 ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion: 43 Resp: 2055
 Ion Ratio Lower Upper
 43 100
 86 8.9 10.2 15.4

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/17/2024
 Supervised By :Semsettin
 Yesilyurt


05/17/2024



#24
 1,1-Dichloroethane
 Concen: 0.725 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19


Tgt Ion: 63 Resp: 3176
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.6 10.9#
 100 0.0 2.6 7.8#





#25
2-Butanone
Concen: 3.354 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 43 Resp: 573
Ion Ratio Lower Upper
43 100
72 28.7 25.3 37.9

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt

05/17/2024



#26
2,2-Dichloropropane
Concen: 0.676 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 77 Resp: 2481
Ion Ratio Lower Upper
77 100
97 10.0 11.3 33.9#





#27
cis-1,2-Dichloroethene
Concen: 0.723 ug/l
RT: 7.494 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 96 Resp: 1768
Ion Ratio Lower Upper
96 100
61 167.9 0.0 268.0
98 57.1 0.0 126.8

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#28
Bromochloromethane
Concen: 0.750 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 49 Resp: 1532
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.6
130 48.0 66.7 100.1#





#29

Tetrahydrofuran

Concen: 3.410 ug/l

RT: 7.847 min Scan# 1000

Delta R.T. 0.011 min

Lab File: VN082157.D


Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024



Tgt Ion:	42	Resp:	374
Ion Ratio	100	Lower	Upper
42	100		
72	35.4	39.3	58.9
71	27.1	37.4	56.0




Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda

05/17/2024

Supervised By :Semsettin
Yesilyurt

05/17/2024



#30

Chloroform

Concen: 0.738 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19


Tgt Ion:	83	Resp:	3087
Ion Ratio	100	Lower	Upper
83	100		
85	65.0	50.7	76.1





#31
Cyclohexane
Concen: 1.727 ug/l
RT: 8.229 min Scan# 1067
Delta R.T. -0.030 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 56 Resp: 758
Ion Ratio Lower Upper
56 100
69 134.7 24.4 36.6
84 121.1 77.0 115.4

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#32
1,1,1-Trichloroethane
Concen: 0.745 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 97 Resp: 2765
Ion Ratio Lower Upper
97 100
99 0.0 50.3 75.5
61 43.0 36.6 55.0





#33
1,2-Dichloroethane-d4
Concen: 52.364 ug/l
RT: 8.582 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 65 Resp: 143969
Ion Ratio Lower Upper
65 100
67 49.4 0.0 103.0

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:114 Resp: 304959
Ion Ratio Lower Upper
114 100
63 24.2 0.0 41.0
88 15.7 0.0 31.4





#35
Dibromofluoromethane
Concen: 50.351 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 113 Resp: 94456
Ion Ratio Lower Upper
113 100
111 103.0 83.4 125.2
192 17.7 14.0 21.0

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#36
1,1-Dichloropropene
Concen: 0.650 ug/l
RT: 8.382 min Scan# 1093
Delta R.T. 0.012 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 75 Resp: 2046
Ion Ratio Lower Upper
75 100
110 0.0 17.8 53.5#
77 0.0 25.0 37.4#





#37
Ethyl Acetate
Concen: 0.714 ug/l
RT: 7.565 min Scan# 914
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion: 43 Resp: 2514
Ion Ratio Lower Upper
43 100
61 0.0 11.3 16.9
70 2.5 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#38
Carbon Tetrachloride
Concen: 0.711 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:117 Resp: 2267
Ion Ratio Lower Upper
117 100
119 63.1 76.5 114.7#
121 26.7 23.9 35.9





#39
Methylcyclohexane
Concen: 0.545 ug/l
RT: 9.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 83 Resp: 1879
Ion Ratio Lower Upper
83 100
55 93.7 53.0 79.6
98 38.1 37.1 55.7

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt

05/17/2024



#40
Benzene
Concen: 0.665 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 78 Resp: 6242
Ion Ratio Lower Upper
78 100
77 21.5 20.0 30.0





#41
Methacrylonitrile
Concen: 0.744 ug/l m
RT: 7.788 min Scan# 990
Delta R.T. 0.012 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 41 Resp: 154
Ion Ratio Lower Upper
41 100
39 45.7 46.6 69.8
67 59.5 67.3 100.9
52 11.4 27.3 40.9

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#42
1,2-Dichloroethane
Concen: 0.699 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 62 Resp: 2523
Ion Ratio Lower Upper
62 100
98 0.0 0.0 19.6





#43
Isopropyl Acetate
Concen: 0.617 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 43 Resp: 3849
Ion Ratio Lower Upper
43 100
61 0.0 22.8 34.2
87 9.7 12.0 18.0

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#44
Trichloroethene
Concen: 0.764 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:130 Resp: 1629
Ion Ratio Lower Upper
130 100
95 84.9 0.0 195.4





#45
1,2-Dichloropropane
Concen: 0.697 ug/l
RT: 9.629 min Scan# 1305
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 63 Resp: 168
Ion Ratio Lower Upper
63 100
65 35.4 24.1 36.1

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#46
Dibromomethane
Concen: 0.746 ug/l
RT: 9.712 min Scan# 1319
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 93 Resp: 1186
Ion Ratio Lower Upper
93 100
95 59.0 64.8 97.2#
174 73.4 73.4 110.0





#47
Bromodichloromethane
Concen: 0.640 ug/l
RT: 9.888 min Scan# 1147
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 83 Resp: 2147
Ion Ratio Lower Upper
83 100
85 64.0 51.9 77.9
127 0.0 7.5 11.3

Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda

05/17/2024
Supervised By :Semsettin
Yesilyurt



05/17/2024




#48
Methyl methacrylate
Concen: 0.679 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19




Tgt Ion: 41 Resp: 2019
Ion Ratio Lower Upper
41 100
69 67.9 77.9 116.9#
39 77.9 50.9 76.3#





#49
1,4-Dioxane
Concen: 13.971 ug/l
RT: 9.700 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 88 Resp: 582
Ion Ratio Lower Upper
88 100
43 0.0 19.0 28.6
58 39.0 50.6 76.0

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#50
Toluene-d8
Concen: 48.701 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 98 Resp: 350270
Ion Ratio Lower Upper
98 100
100 65.0 51.8 77.6





#51
4-Methyl-2-Pentanone
Concen: 3.230 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 43 Resp: 11450
Ion Ratio Lower Upper
43 100
58 33.0 33.2 49.8

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#52
Toluene
Concen: 0.698 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 92 Resp: 3932
Ion Ratio Lower Upper
92 100
91 174.1 136.7 205.1





#53
t-1,3-Dichloropropene
Concen: 0.659 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion: 75 Resp: 2280
Ion Ratio Lower Upper
75 100
77 22.3 25.4 38.0#

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#54
cis-1,3-Dichloropropene
Concen: 0.690 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 75 Resp: 2536
Ion Ratio Lower Upper
75 100
77 23.9 25.0 37.4#
39 56.4 34.8 52.2#





#55
1,1,2-Trichloroethane
Concen: 0.724 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion: 97 Resp: 148
Ion Ratio Lower Upper
97 100
83 111.1 70.4 105.6
85 44.5 45.6 68.4
99 53.5 50.9 76.3

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#56
Ethyl methacrylate
Concen: 0.562 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 69 Resp: 1951
Ion Ratio Lower Upper
69 100
41 96.4 46.6 70.0#
39 54.3 26.5 39.7#





#57

1,3-Dichloropropane

Concen: 0.715 ug/l

RT: 11.170 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion: 76 Resp: 2738
 Ion Ratio Lower Upper
 76 100
 78 32.1 25.8 38.6

Manual Integrations





APPROVED

Reviewed By :Mahesh
 Dadoda

05/17/2024

Supervised By :Semsettin
 Yesilyurt

05/17/2024



#58

2-Chloroethyl Vinyl ether

Concen: 3.146 ug/l




RT: 10.165 min Scan# 1396


Delta R.T. 0.006 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19


Tgt Ion: 63 Resp: 5637
 Ion Ratio Lower Upper
 63 100
 106 26.0 24.6 36.8





#59
2-Hexanone
Concen: 3.126 ug/l
RT: 11.200 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion: 43 Resp: 811
Ion Ratio Lower Upper
43 100
58 46.8 28.3 84.9

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#60
Dibromochloromethane
Concen: 0.642 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:129 Resp: 1449
Ion Ratio Lower Upper
129 100
127 84.5 38.3 114.8





#61
1,2-Dibromoethane
Concen: 0.675 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion:107 Resp: 1418
Ion Ratio Lower Upper
107 100
109 72.4 75.8 113.6

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#62
4-Bromofluorobenzene
Concen: 43.719 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion: 95 Resp: 112231
Ion Ratio Lower Upper
95 100
174 76.2 0.0 160.6
176 75.1 0.0 152.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion:117 Resp: 264950
Ion Ratio Lower Upper
117 100
82 63.4 44.6 66.8
119 32.8 25.9 38.9

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#64
Tetrachloroethene
Concen: 0.767 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:164 Resp: 1507
Ion Ratio Lower Upper
164 100
166 120.5 97.6 146.4
129 104.7 73.7 110.5
131 72.6 75.8 113.6#





#65
Chlorobenzene
Concen: 0.718 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion:112 Resp: 420
Ion Ratio Lower Upper
112 100
114 30.9 25.7 38.5

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 0.579 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion:131 Resp: 1116
Ion Ratio Lower Upper
131 100
133 110.4 47.5 142.6
119 0.0 32.9 98.7#





#67

Ethyl Benzene

Concen: 0.612 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion: 91 Resp: 640
 Ion Ratio Lower Upper
 91 100
 106 33.5 25.3 37.9





Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda

05/17/2024

Supervised By :Semsettin
 Yesilyurt

05/17/2024



#68

m/p-Xylenes

Concen: 1.253 ug/l




RT: 12.070 min Scan# 1720


Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19


Tgt Ion:106 Resp: 4809
 Ion Ratio Lower Upper
 106 100
 91 212.5 159.4 239.0





#69
o-Xylene
Concen: 0.646 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024






Tgt Ion:106 Resp: 240
Ion Ratio Lower Upper
106 100
91 219.0 104.7 314.1

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#70
Styrene
Concen: 0.536 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:104 Resp: 3312
Ion Ratio Lower Upper
104 100
78 52.5 38.6 58.0
103 56.4 43.8 65.6





#71
Bromoform
Concen: 0.574 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion:	173	Resp:	79
Ion Ratio	Lower	Upper	
173	100		
175	50.3	24.6	74.0
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion:	152	Resp:	107860
Ion Ratio	Lower	Upper	
152	100		
115	61.7	30.4	91.3
150	160.2	0.0	353.2





#73

Isopropylbenzene

Concen: 0.710 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:105 Resp: 6080
 Ion Ratio Lower Upper
 105 100
 120 25.5 13.4 40.1

Manual Integrations





APPROVED

Reviewed By :Mahesh
 Dadoda

05/17/2024

Supervised By :Semsettin
 Yesilyurt

05/17/2024



#74

N-ethyl acetate

Concen: 0.649 ug/l




RT: 12.500 min Scan# 1793


Delta R.T. 0.006 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19


Tgt Ion: 43 Resp: 3059
 Ion Ratio Lower Upper
 43 100
 70 35.2 39.0 58.4#
 55 25.8 21.3 31.9
 61 21.0 20.5 30.7





#75
1,1,2,2-Tetrachloroethane
Concen: 0.885 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion: 83 Resp: 2180
Ion Ratio Lower Upper
83 100
131 0.0 5.5 16.4
85 48.2 32.9 98.6

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#76
1,2,3-Trichloropropane
Concen: 0.853 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 75 Resp: 1937
Ion Ratio Lower Upper
75 100
77 174.5 90.5 271.3





#77
Bromobenzene
Concen: 0.898 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion:156 Resp: 183
Ion Ratio Lower Upper
156 100
77 184.5 100.2 300.6
158 70.9 49.1 147.3

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#78
n-propylbenzene
Concen: 0.668 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 91 Resp: 6783
Ion Ratio Lower Upper
91 100
120 17.9 11.7 35.0





#79
2-Chlorotoluene
Concen: 0.700 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024





Tgt Ion: 91 Resp: 446
Ion Ratio Lower Upper
91 100
126 31.9 17.5 52.5

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#80
1,3,5-Trimethylbenzene
Concen: 0.619 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19


Tgt Ion:105 Resp: 4290
Ion Ratio Lower Upper
105 100
120 48.4 24.8 74.3





#81
trans-1,4-Dichloro-2-butene
Concen: 0.455 ug/l m
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT2-2024






Tgt Ion: 75 Resp: 442
Ion Ratio Lower Upper
75 100
53 0.0 72.9 109.3
89 0.0 39.1 58.7

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#82
4-Chlorotoluene
Concen: 0.687 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion: 91 Resp: 4281
Ion Ratio Lower Upper
91 100
126 27.4 17.4 52.3





#83

tert-Butylbenzene

Concen: 0.670 ug/l

RT: 13.441 min Scan# 1953

Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:	119	Resp:	4060
Ion	Ratio	Lower	Upper
119	100		
91	72.6	32.6	97.8
134	25.8	12.4	37.0





Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda

05/17/2024

Supervised By :Semsettin
Yesilyurt

05/17/2024



#84

1,2,4-Trimethylbenzene

Concen: 0.608 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Tgt Ion:	105	Resp:	4310
Ion	Ratio	Lower	Upper
105	100		
120	40.2	22.9	68.7





#85
 sec-Butylbenzene
 Concen: 0.648 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT2-2024





Tgt Ion:105 Resp: 544
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.2 30.4

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/17/2024
 Supervised By :Semsettin
 Yesilyurt


05/17/2024



#86
 p-Isopropyltoluene
 Concen: 0.561 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN082157.D
 Acq: 16 May 2024 13:19

Tgt Ion:119 Resp: 3832
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.2 39.5
 91 36.2 11.8 35.3#





#87
1,3-Dichlorobenzene
Concen: 0.695 ug/l
RT: 13.729 min Scan# 2003
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion:146 Resp: 257
Ion Ratio Lower Upper
146 100
111 48.6 20.8 62.5
148 59.3 32.1 96.3

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#88
1,4-Dichlorobenzene
Concen: 0.719 ug/l m
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion:146 Resp: 2714
Ion Ratio Lower Upper
146 100
111 46.0 19.3 57.9
148 56.2 31.3 93.9





#89
n-Butylbenzene
Concen: 0.597 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.006 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Instrument : MSVOA_N
ClientSampleId : MDL-WATER-03-QT2-2024





Tgt Ion: 91 Resp: 366
Ion Ratio Lower Upper
91 100
92 45.0 26.2 78.5
134 24.1 13.2 39.5

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/17/2024
Supervised By :Semsettin
Yesilyurt


05/17/2024



#90
Hexachloroethane
Concen: 0.665 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN082157.D
Acq: 16 May 2024 13:19

Tgt Ion:117 Resp: 841
Ion Ratio Lower Upper
117 100
201 48.9 33.3 99.8





#91

1,2-Dichlorobenzene

Concen: 0.692 ug/l

RT: 14.111 min Scan# 2066

Delta R.T. 0.006 min

Lab File: VN082157.D


Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024



Tgt Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.3	64.0
148	76.0	31.8	95.4

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda

05/17/2024

Supervised By :Semsettin
Yesilyurt

05/17/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.339 ug/l




RT: 14.723 min Scan# 2171


Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Tgt Ion	Ratio	Lower	Upper
75	100		
155	0.0	42.6	128.0#
157	98.8	53.0	159.0





#93

1,2,4-Trichlorobenzene

Concen: 0.625 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.006 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:180	Resp:	122
Ion Ratio	Lower	Upper
180	100	
182	88.6	46.7 140.0
145	23.2	16.2 48.6

Manual Integrations





APPROVED

Reviewed By :Mahesh
Dadoda

05/17/2024

Supervised By :Semsettin
Yesilyurt

05/17/2024



#94

Hexachlorobutadiene

Concen: 0.715 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Tgt Ion:225	Resp:	616
Ion Ratio	Lower	Upper
225	100	
223	46.3	31.1 93.2
227	45.0	30.6 91.6





#95

Naphthalene

Concen: 0.542 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Instrument :

MSVOA_N

ClientSampleId :

MDL-WATER-03-QT2-2024

Tgt Ion:128	Resp:	3330
Ion Ratio	Lower	Upper
128	100	
127	7.7	10.1 15.1
129	3.8	8.9 13.3





Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda

05/17/2024

Supervised By :Semsettin
Yesilyurt

05/17/2024



#96

1,2,3-Trichlorobenzene

Concen: 0.586 ug/l

RT: 15.847 min Scan# 2362

Delta R.T. 0.006 min

Lab File: VN082157.D

Acq: 16 May 2024 13:19

Tgt Ion:180	Resp:	1152
Ion Ratio	Lower	Upper
180	100	
182	75.7	46.0 137.8
145	23.4	17.1 51.3

