

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051625\
 Data File : VN086664.D
 Acq On : 16 May 2025 13:20
 Operator : JC\MD
 Sample : VN0516WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0516WBL01

Quant Time: May 16 23:03:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Apr 24 04:42:24 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	29934	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	171805	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	155509	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	82632	28.632	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.433%
60) 4-Bromofluorobenzene	12.847	95	81087	26.530	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.433%
63) Toluene-d8	10.565	98	254603	29.760	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051625\
Data File : VN086664.D
Acq On : 16 May 2025 13:20
Operator : JC\MD
Sample : VN0516WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0516WBL01

Quant Time: May 16 23:03:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Apr 24 04:42:24 2025
Response via : Initial Calibration

