


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051625\
Data File : VN086660.D
Acq On : 16 May 2025 09:17
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042325W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu Apr 24 04:42:24 2025



AutoFind: Averaged scan 1845 to 1860; Bkg corrected with scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	3188	PASS
75	95	30	60	49.9	8949	PASS
95	95	100	100	100.0	17923	PASS
96	95	5	9	7.0	1246	PASS
173	174	0.00	2	1.0	129	PASS
174	95	50	100	72.6	13020	PASS
175	174	5	9	6.8	879	PASS
176	174	95	101	95.5	12438	PASS
177	176	5	9	6.5	813	PASS