

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN051719\
 Data File : VN055626.D
 Acq On : 17 May 2019 21:29
 Operator : JC/SP
 Sample : K2823-04
 Misc : 5.62g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11-B2

Quant Time: May 20 09:37:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	332253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	475045	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	405303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	178434	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	158356	43.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.96%	
35) Dibromofluoromethane	7.59	113	143474	47.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.08%	
50) Toluene-d8	10.09	98	552804	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.40	95	182966	46.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055626.D
Acq On : 17 May 2019 21:29
Operator : JC/SP
Sample : K2823-04
Misc : 5.62µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11-B2

Quant Time: May 20 09:37:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

