

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055628.D
 Acq On : 17 May 2019 22:16
 Operator : JC/SP
 Sample : K2823-06ME
 Misc : 5.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 12-B2ME

Quant Time: May 18 04:16:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	356623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	508707	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	441184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	219864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	168192	43.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.04%	
35) Dibromofluoromethane	7.59	113	151853	46.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.98%	
50) Toluene-d8	10.09	98	602627	49.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.64%	
62) 4-Bromofluorobenzene	12.40	95	213117	50.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.02%	
Target Compounds						
					Qvalue	
31) Cyclohexane	7.63	56	153924	26.548	ug/l #	52
39) Methylcyclohexane	9.08	83	224340	41.032	ug/l	90
73) Isopropylbenzene	12.25	105	61443	1.653	ug/l	100
78) n-propylbenzene	12.59	91	429408	21.993	ug/l	100
85) sec-Butylbenzene	13.17	105	159118	9.075	ug/l #	49
89) n-Butylbenzene	13.62	91	427560	35.439	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055628.D
Acq On : 17 May 2019 22:16
Operator : JC/SP
Sample : K2823-06ME
Misc : 5.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
12-B2ME

Quant Time: May 18 04:16:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

