

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055636.D
 Acq On : 18 May 2019 2:35
 Operator : JC/SP
 Sample : VN0517WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0517WBL02

Quant Time: May 18 04:37:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	414033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	584925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	488343	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	190109	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	191204	42.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.26%	
35) Dibromofluoromethane	7.59	113	176996	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
50) Toluene-d8	10.09	98	676677	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
62) 4-Bromofluorobenzene	12.40	95	211810	43.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055636.D
Acq On : 18 May 2019 2:35
Operator : JC/SP
Sample : VN0517WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0517WBL02

Quant Time: May 18 04:37:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

