

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055648.D
 Acq On : 18 May 2019 7:17
 Operator : JC/SP
 Sample : K2643-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-051619-B

Quant Time: May 20 02:55:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	401246	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	578032	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	499352	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	214769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	191949	43.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.28%	
35) Dibromofluoromethane	7.59	113	177985	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	10.09	98	669783	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	12.40	95	225504	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
Target Compounds						
16) Acetone	3.82	43	10497	5.848	ug/l	96
20) Methylene Chloride	4.56	84	146197	35.689	ug/l	99
43) Isopropyl Acetate	8.17	43	13742	1.699	ug/l #	91
95) Naphthalene	15.13	128	8167	2.470	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055648.D
Acq On : 18 May 2019 7:17
Operator : JC/SP
Sample : K2643-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-051619-B

Quant Time: May 20 02:55:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

