

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055651.D
 Acq On : 18 May 2019 8:27
 Operator : JC/SP
 Sample : K2880-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 339-RBR-251608

Quant Time: May 20 03:05:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	388707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	558005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	478927	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	211283	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	184258	43.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.48%	
35) Dibromofluoromethane	7.59	113	172247	48.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.18%	
50) Toluene-d8	10.09	98	644920	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
62) 4-Bromofluorobenzene	12.40	95	218969	47.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.62%	
Target Compounds						
16) Acetone	3.82	43	15614	8.979	ug/l	97
20) Methylene Chloride	4.56	84	126279	31.821	ug/l	98
67) Ethyl Benzene	11.51	91	24149	1.399	ug/l	98
68) m/p-Xylenes	11.62	106	42004	6.446	ug/l	98
69) o-Xylene	11.95	106	19092	2.995	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	150506	10.522	ug/l	99
95) Naphthalene	15.13	128	37352	5.044	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055651.D
Acq On : 18 May 2019 8:27
Operator : JC/SP
Sample : K2880-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 59 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
339-RBR-251608

Quant Time: May 20 03:05:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

