

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055655.D
 Acq On : 18 May 2019 10:01
 Operator : JC/SP
 Sample : K2899-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OIL-WATER-SEPERATOR(OWS)-4

Quant Time: May 20 03:53:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	399672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	579651	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	499607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	220672	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	190043	43.38	ug/l	0.00
Spiked Amount			Recovery	=	86.76%	
35) Dibromofluoromethane	7.59	113	177771	48.28	ug/l	0.00
Spiked Amount			Recovery	=	96.56%	
50) Toluene-d8	10.09	98	667452	47.94	ug/l	0.00
Spiked Amount			Recovery	=	95.88%	
62) 4-Bromofluorobenzene	12.40	95	227240	47.27	ug/l	0.00
Spiked Amount			Recovery	=	94.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	78755	44.048	ug/l	96
20) Methylene Chloride	4.56	84	5762	1.412	ug/l	95
25) 2-Butanone	6.84	43	27602	11.041	ug/l	97
43) Isopropyl Acetate	8.18	43	18976	2.339	ug/l #	77
51) 4-Methyl-2-Pentanone	9.99	43	30859	6.485	ug/l	88
52) Toluene	10.16	92	138362	14.512	ug/l	99
86) p-Isopropyltoluene	13.29	119	186389	12.251	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055655.D
Acq On : 18 May 2019 10:01
Operator : JC/SP
Sample : K2899-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 63 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OIL-WATER-SEPERATOR(OWS)-4

Quant Time: May 20 03:53:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

