

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055633.D
 Acq On : 18 May 2019 00:14
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:29:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.385	0.346	10.1	100	0.00
3 P	Chloromethane	0.513	0.451	12.1	98	0.00
4 C	Vinyl Chloride	0.556	0.523	5.9#	104	0.00
5 T	Bromomethane	0.354	0.344	2.8	106	0.00
6 T	Chloroethane	0.309	0.298	3.6	102	0.00
7 T	Trichlorofluoromethane	0.712	0.684	3.9	104	0.00
8 T	Diethyl Ether	0.286	0.280	2.1	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.434	0.410	5.5	101	0.00
10 T	Methyl Iodide	0.626	0.600	4.2	99	0.00
11 T	Tert butyl alcohol	0.102	0.106	-3.9	108	0.00
12 CM	1,1-Dichloroethene	0.457	0.430	5.9#	103	0.00
13 T	Acrolein	0.055	0.050	9.1	103	0.00
14 T	Allyl chloride	0.737	0.699	5.2	99	0.00
15 T	Acrylonitrile	0.227	0.231	-1.8	105	0.00
16 T	Acetone	0.224	0.189	15.6	89	0.00
17 T	Carbon Disulfide	1.269	1.170	7.8	101	0.00
18 T	Methyl Acetate	0.480	0.483	-0.6	107	0.00
19 T	Methyl tert-butyl Ether	1.511	1.467	2.9	104	0.00
20 T	Methylene Chloride	0.510	0.478	6.3	102	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.466	4.1	103	0.00
22 T	Diisopropyl ether	1.461	1.404	3.9	103	0.00
23 T	Vinyl Acetate	1.239	1.217	1.8	102	0.00
24 P	1,1-Dichloroethane	0.856	0.824	3.7	102	0.00
25 T	2-Butanone	0.313	0.308	1.6	102	0.00
26 T	2,2-Dichloropropane	0.760	0.548	27.9#	78	0.00
27 T	cis-1,2-Dichloroethene	0.566	0.548	3.2	105	0.00
28 T	Bromochloromethane	0.373	0.374	-0.3	102	0.00
29 T	Tetrahydrofuran	0.207	0.207	0.0	106	0.00
30 C	Chloroform	0.879	0.838	4.7#	104	0.00
31 T	Cyclohexane	0.813	0.743	8.6	101	0.00
32 T	1,1,1-Trichloroethane	0.789	0.739	6.3	102	0.00
33 S	1,2-Dichloroethane-d4	0.548	0.544	0.7	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.318	0.321	-0.9	105	0.00
36 T	1,1-Dichloropropene	0.446	0.414	7.2	101	0.00
37 T	Ethyl Acetate	0.402	0.417	-3.7	104	0.00
38 T	Carbon Tetrachloride	0.476	0.440	7.6	103	0.00
39 T	Methylcyclohexane	0.537	0.478	11.0	98	0.00
40 TM	Benzene	1.311	1.265	3.5	104	0.00
41 T	Methacrylonitrile	0.172	0.176	-2.3	98	0.00
42 TM	1,2-Dichloroethane	0.425	0.414	2.6	104	0.00
43 T	Isopropyl Acetate	0.700	0.686	2.0	105	0.00
44 TM	Trichloroethene	0.370	0.368	0.5	107	0.00
45 C	1,2-Dichloropropane	0.345	0.330	4.3#	103	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055633.D
 Acq On : 18 May 2019 00:14
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:29:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.226	0.221	2.2	104	0.00
47 T	Bromodichloromethane	0.474	0.452	4.6	104	0.00
48 T	Methyl methacrylate	0.313	0.312	0.3	102	0.00
49 T	1,4-Dioxane	0.006	0.007	-16.7	105	0.00
50 S	Toluene-d8	1.201	1.195	0.5	104	0.00
51 T	4-Methyl-2-Pentanone	0.410	0.407	0.7	107	0.00
52 CM	Toluene	0.822	0.792	3.6#	104	0.00
53 T	t-1,3-Dichloropropene	0.510	0.486	4.7	100	0.00
54 T	cis-1,3-Dichloropropene	0.559	0.528	5.5	99	0.00
55 T	1,1,2-Trichloroethane	0.339	0.326	3.8	106	0.00
56 T	Ethyl methacrylate	0.529	0.504	4.7	103	0.00
57 T	1,3-Dichloropropane	0.536	0.527	1.7	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.225	0.219	2.7	101	0.00
59 T	2-Hexanone	0.289	0.288	0.3	104	0.00
60 T	Dibromochloromethane	0.384	0.386	-0.5	107	0.00
61 T	1,2-Dibromoethane	0.339	0.339	0.0	106	0.00
62 S	4-Bromofluorobenzene	0.415	0.412	0.7	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.394	0.391	0.8	107	0.00
65 PM	Chlorobenzene	1.011	0.978	3.3	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.394	3.2	106	0.00
67 C	Ethyl Benzene	1.802	1.705	5.4#	103	0.00
68 T	m/p-Xylenes	0.680	0.646	5.0	102	0.00
69 T	o-Xylene	0.665	0.629	5.4	103	0.00
70 T	Styrene	1.106	1.087	1.7	103	0.00
71 P	Bromoform	0.345	0.352	-2.0	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	4.076	3.627	11.0	103	0.00
74 T	N-amyl acetate	1.404	1.388	1.1	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.259	1.140	9.5	104	0.00
76 T	1,2,3-Trichloropropane	1.020	0.946	7.3	105	0.00
77 T	Bromobenzene	1.027	0.976	5.0	105	0.00
78 T	n-propylbenzene	4.440	3.966	10.7	100	0.00
79 T	2-Chlorotoluene	2.696	2.366	12.2	100	0.00
80 T	1,3,5-Trimethylbenzene	3.428	2.996	12.6	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.356	10.3	95	0.00
82 T	4-Chlorotoluene	2.616	2.387	8.8	100	0.00
83 T	tert-Butylbenzene	3.002	2.597	13.5	101	0.00
84 T	1,2,4-Trimethylbenzene	3.374	2.995	11.2	100	0.00
85 T	sec-Butylbenzene	3.818	3.349	12.3	99	0.00
86 T	p-Isopropyltoluene	3.509	3.081	12.2	99	0.00
87 T	1,3-Dichlorobenzene	1.698	1.613	5.0	101	0.00
88 T	1,4-Dichlorobenzene	1.621	1.585	2.2	102	0.00
89 T	n-Butylbenzene	2.744	2.483	9.5	96	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055633.D
Acq On : 18 May 2019 00:14
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 18 04:29:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.672	0.617	8.2	102	0.00
91 T	1,2-Dichlorobenzene	1.692	1.577	6.8	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.221	2.6	100	0.00
93 T	1,2,4-Trichlorobenzene	0.862	0.925	-7.3	101	0.00
94 T	Hexachlorobutadiene	0.595	0.513	13.8	95	0.00
95 T	Naphthalene	2.320	2.635	-13.6	106	0.00
96 T	1,2,3-Trichlorobenzene	0.901	0.932	-3.4	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6