

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055635.D
 Acq On : 18 May 2019 1:48
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:34:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.385	0.349	9.4	94	0.00
3 P	Chloromethane	0.513	0.484	5.7	98	0.00
4 C	Vinyl Chloride	0.556	0.541	2.7#	100	0.00
5 T	Bromomethane	0.354	0.357	-0.8	102	0.00
6 T	Chloroethane	0.309	0.311	-0.6	99	0.00
7 T	Trichlorofluoromethane	0.712	0.698	2.0	98	0.00
8 T	Diethyl Ether	0.286	0.289	-1.0	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.434	0.407	6.2	94	0.00
10 T	Methyl Iodide	0.626	0.625	0.2	96	0.00
11 T	Tert butyl alcohol	0.102	0.109	-6.9	103	0.00
12 CM	1,1-Dichloroethene	0.457	0.439	3.9#	98	0.00
13 T	Acrolein	0.055	0.050	9.1	97	0.00
14 T	Allyl chloride	0.737	0.719	2.4	94	0.00
15 T	Acrylonitrile	0.227	0.240	-5.7	102	0.00
16 T	Acetone	0.224	0.198	11.6	87	0.00
17 T	Carbon Disulfide	1.269	1.201	5.4	96	0.00
18 T	Methyl Acetate	0.480	0.492	-2.5	102	0.00
19 T	Methyl tert-butyl Ether	1.511	1.520	-0.6	100	0.00
20 T	Methylene Chloride	0.510	0.502	1.6	100	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.477	1.9	98	0.00
22 T	Diisopropyl ether	1.461	1.431	2.1	98	0.00
23 T	Vinyl Acetate	1.239	1.240	-0.1	97	0.00
24 P	1,1-Dichloroethane	0.856	0.855	0.1	98	0.00
25 T	2-Butanone	0.313	0.318	-1.6	98	0.00
26 T	2,2-Dichloropropane	0.760	0.535	29.6#	71	0.00
27 T	cis-1,2-Dichloroethene	0.566	0.559	1.2	99	0.00
28 T	Bromochloromethane	0.373	0.384	-2.9	98	0.00
29 T	Tetrahydrofuran	0.207	0.213	-2.9	102	0.00
30 C	Chloroform	0.879	0.866	1.5#	100	0.00
31 T	Cyclohexane	0.813	0.745	8.4	95	0.00
32 T	1,1,1-Trichloroethane	0.789	0.770	2.4	99	0.00
33 S	1,2-Dichloroethane-d4	0.548	0.558	-1.8	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.318	0.325	-2.2	101	0.00
36 T	1,1-Dichloropropene	0.446	0.419	6.1	97	0.00
37 T	Ethyl Acetate	0.402	0.421	-4.7	100	0.00
38 T	Carbon Tetrachloride	0.476	0.447	6.1	99	0.00
39 T	Methylcyclohexane	0.537	0.474	11.7	91	0.00
40 TM	Benzene	1.311	1.281	2.3	100	0.00
41 T	Methacrylonitrile	0.172	0.199	-15.7	105	0.00
42 TM	1,2-Dichloroethane	0.425	0.423	0.5	100	0.00
43 T	Isopropyl Acetate	0.700	0.691	1.3	100	0.00
44 TM	Trichloroethene	0.370	0.368	0.5	101	0.00
45 C	1,2-Dichloropropane	0.345	0.337	2.3#	99	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055635.D
 Acq On : 18 May 2019 1:48
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 18 04:34:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.226	0.223	1.3	100	0.00
47 T	Bromodichloromethane	0.474	0.456	3.8	99	0.00
48 T	Methyl methacrylate	0.313	0.317	-1.3	98	0.00
49 T	1,4-Dioxane	0.006	0.007	-16.7	103	0.00
50 S	Toluene-d8	1.201	1.206	-0.4	100	0.00
51 T	4-Methyl-2-Pentanone	0.410	0.412	-0.5	102	0.00
52 CM	Toluene	0.822	0.805	2.1#	100	0.00
53 T	t-1,3-Dichloropropene	0.510	0.490	3.9	95	0.00
54 T	cis-1,3-Dichloropropene	0.559	0.527	5.7	94	0.00
55 T	1,1,2-Trichloroethane	0.339	0.334	1.5	102	0.00
56 T	Ethyl methacrylate	0.529	0.515	2.6	100	0.00
57 T	1,3-Dichloropropane	0.536	0.533	0.6	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.225	0.222	1.3	97	0.00
59 T	2-Hexanone	0.289	0.294	-1.7	100	0.00
60 T	Dibromochloromethane	0.384	0.389	-1.3	102	0.00
61 T	1,2-Dibromoethane	0.339	0.341	-0.6	101	0.00
62 S	4-Bromofluorobenzene	0.415	0.425	-2.4	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.394	0.395	-0.3	104	0.00
65 PM	Chlorobenzene	1.011	0.993	1.8	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.394	3.2	102	0.00
67 C	Ethyl Benzene	1.802	1.718	4.7#	100	0.00
68 T	m/p-Xylenes	0.680	0.656	3.5	100	0.00
69 T	o-Xylene	0.665	0.642	3.5	101	0.00
70 T	Styrene	1.106	1.103	0.3	101	0.00
71 P	Bromoform	0.345	0.354	-2.6	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.076	3.615	11.3	100	0.00
74 T	N-amyl acetate	1.404	1.383	1.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.259	1.143	9.2	102	0.00
76 T	1,2,3-Trichloropropane	1.020	0.945	7.4	103	0.00
77 T	Bromobenzene	1.027	0.992	3.4	104	0.00
78 T	n-propylbenzene	4.440	3.957	10.9	97	0.00
79 T	2-Chlorotoluene	2.696	2.409	10.6	100	0.00
80 T	1,3,5-Trimethylbenzene	3.428	3.019	11.9	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.339	14.6	88	0.00
82 T	4-Chlorotoluene	2.616	2.421	7.5	99	0.00
83 T	tert-Butylbenzene	3.002	2.596	13.5	98	0.00
84 T	1,2,4-Trimethylbenzene	3.374	2.982	11.6	98	0.00
85 T	sec-Butylbenzene	3.818	3.368	11.8	98	0.00
86 T	p-Isopropyltoluene	3.509	3.093	11.9	97	0.00
87 T	1,3-Dichlorobenzene	1.698	1.643	3.2	101	0.00
88 T	1,4-Dichlorobenzene	1.621	1.615	0.4	101	0.00
89 T	n-Butylbenzene	2.744	2.496	9.0	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055635.D
Acq On : 18 May 2019 1:48
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 18 04:34:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.672	0.606	9.8	98	0.00
91 T	1,2-Dichlorobenzene	1.692	1.625	4.0	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.228	-0.4	101	0.00
93 T	1,2,4-Trichlorobenzene	0.862	0.945	-9.6	100	0.00
94 T	Hexachlorobutadiene	0.595	0.531	10.8	96	0.00
95 T	Naphthalene	2.320	2.675	-15.3	105	0.00
96 T	1,2,3-Trichlorobenzene	0.901	0.944	-4.8	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6