

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051719\
 Data File : VN055613.D
 Acq On : 17 May 2019 16:22
 Operator : JC/SP
 Sample : K2918-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 18 03:36:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	393679	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	564367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	471490	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	181532	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	189268	43.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.72%	
35) Dibromofluoromethane	7.59	113	173625	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
50) Toluene-d8	10.09	98	652269	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
62) 4-Bromofluorobenzene	12.40	95	207572	44.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.68%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051719\
Data File : VN055613.D
Acq On : 17 May 2019 16:22
Operator : JC/SP
Sample : K2918-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 18 03:36:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN055613.D

Acq: 17 May 2019 16:22

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion:168 Resp: 393679

Ion Ratio Lower Upper

168 100

99 49.0 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

150000



100000

50000

0

7.60 7.70 7.80

Time-->



#33

1,2-Dichloroethane-d4

Concen: 43.858 ug/l

RT: 8.03 min Scan# 1946

Delta R.T. 0.00 min

Lab File: VN055613.D


Acq: 17 May 2019 16:22

Tgt Ion: 65 Resp: 189268

Ion Ratio Lower Upper

65 100

67 54.1 0.0 108.6



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.03

80000

60000


40000


20000

0

7.90 8.00 8.10 8.20

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2120
 Delta R.T. 0.00 min
 Lab File: VN055613.D
 Acq: 17 May 2019 16:22

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Tot Ion:	114	Resp:	564367
Ion	Ratio	Lower	Upper
114	100		
63	19.2	0.0	38.2
88	15.4	0.0	31.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 48.429 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. -0.00 min
 Lab File: VN055613.D
 Acq: 17 May 2019 16:22

Tgt Ion:	113	Resp:	173625
Ion	Ratio	Lower	Upper
113	100		
111	102.7	82.5	123.7
192	23.2	18.7	28.1



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.117 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN055613.D

Acq: 17 May 2019 16:22

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion: 98 Resp: 652269

Ion Ratio Lower Upper

98 100

100 65.0 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

400000


300000

200000


100000

0

Time-->



Time-->



#62

4-Bromofluorobenzene

Concen: 44.343 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN055613.D

Acq: 17 May 2019 16:22


Tgt Ion: 95 Resp: 207572

Ion Ratio Lower Upper

95 100

174 91.1 0.0 180.6

176 88.0 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN055613.D
Acq: 17 May 2019 16:22

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tot Ion:117 Resp: 471490
Ion Ratio Lower Upper
117 100
82 52.3 42.3 63.5
119 32.2 25.4 38.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN055613.D
Acq: 17 May 2019 16:22

Tgt Ion:152 Resp: 181532
Ion Ratio Lower Upper
152 100
115 56.7 28.1 84.2
150 157.6 0.0 348.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

