

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067092.D
 Acq On : 17 May 2021 11:24
 Operator : JC/MD
 Sample : VN0517WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0517WBL01

Quant Time: May 18 04:04:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.597	168	668532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	954336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.357	117	820217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	303013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	313662	43.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.86%
35) Dibromofluoromethane	7.519	113	289990	46.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.56%
50) Toluene-d8	10.038	98	1120998	47.20	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.40%
62) 4-Bromofluorobenzene	12.357	95	345006	43.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.74%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067092.D
Acq On : 17 May 2021 11:24
Operator : JC/MD
Sample : VN0517WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0517WBL01

Quant Time: May 18 04:04:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

