

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067097.D
 Acq On : 17 May 2021 13:38
 Operator : JC/MD
 Sample : M2387-02
 Misc : 1.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30782

Quant Time: May 18 04:06:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	667727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	944115	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	764750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.299	152	382330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	310651	43.06	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	86.12%		
35) Dibromofluoromethane	7.516	113	286780	46.26	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.52%		
50) Toluene-d8	10.035	98	1120661	47.70	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.40%		
62) 4-Bromofluorobenzene	12.360	95	328094	41.69	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	83.38%		
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.018	83	37203	3.69	ug/l	95
52) Toluene	10.099	92	1482954	87.76	ug/l	98
67) Ethyl Benzene	11.464	91	185399	6.45	ug/l	100
68) m/p-Xylenes	11.572	106	243352	21.92	ug/l	99
69) o-Xylene	11.902	106	192605	17.24	ug/l	97
73) Isopropylbenzene	12.205	105	86633	2.71	ug/l	98
78) n-propylbenzene	12.548	91	357019	11.05	ug/l	99
80) 1,3,5-Trimethylbenzene	12.687	105	595242	22.34	ug/l	99
84) 1,2,4-Trimethylbenzene	12.996	105	1615954	65.78	ug/l	99
85) sec-Butylbenzene	13.127	105	146050	4.86	ug/l	78
86) p-Isopropyltoluene	13.245	119	95732	3.67	ug/l	98
89) n-Butylbenzene	13.570	91	193449	9.78	ug/l #	67
95) Naphthalene	15.088	128	195687	14.52	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067097.D
Acq On : 17 May 2021 13:38
Operator : JC/MD
Sample : M2387-02
Misc : 1.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30782

Quant Time: May 18 04:06:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

