

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067102.D
 Acq On : 17 May 2021 15:31
 Operator : JC/MD
 Sample : M2405-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BUR-1022

Quant Time: May 18 04:08:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.600	168	693909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	983620	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	841421	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.304	152	335617	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	319666	42.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.28%
35) Dibromofluoromethane	7.522	113	296372	45.89	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.78%
50) Toluene-d8	10.037	98	1155120	47.19	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.38%
62) 4-Bromofluorobenzene	12.360	95	346432	42.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.50%
Target Compounds						
17) Carbon Disulfide	4.011	76	12860	0.74	ug/l	96
95) Naphthalene	15.104	128	87658	9.00	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067102.D
Acq On : 17 May 2021 15:31
Operator : JC/MD
Sample : M2405-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BUR-1022

Quant Time: May 18 04:08:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

