

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067107.D
 Acq On : 17 May 2021 17:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 18 04:10:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.600	168	651124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	913580	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	795377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.302	152	338547	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.959	65	295151	41.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.92%
35) Dibromofluoromethane	7.519	113	270859	45.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.30%
50) Toluene-d8	10.038	98	1074070	47.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.48%
62) 4-Bromofluorobenzene	12.357	95	342900	45.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.04%
Target Compounds						
16) Acetone	3.772	43	21633	9.90	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067107.D
Acq On : 17 May 2021 17:24
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 18 04:10:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

