

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067108.D
 Acq On : 17 May 2021 17:47
 Operator : JC/MD
 Sample : M2387-03 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31055

Quant Time: May 18 04:10:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.597	168	661679	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	934753	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	825335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	337514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	309238	43.26	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	86.52%		
35) Dibromofluoromethane	7.522	113	284650	46.38	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.76%		
50) Toluene-d8	10.037	98	1116089	47.98	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.96%		
62) 4-Bromofluorobenzene	12.357	95	353934	45.42	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.84%		
Target Compounds						
16) Acetone	3.775	43	8163	3.67	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067108.D
Acq On : 17 May 2021 17:47
Operator : JC/MD
Sample : M2387-03 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31055

Quant Time: May 18 04:10:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

