

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067110.D
 Acq On : 17 May 2021 18:35
 Operator : JC/MD
 Sample : M2387-09DL 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30794DL

Quant Time: May 18 04:11:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.600	168	633790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	894605	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.357	117	764862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	329290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	301803	44.08	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.16%
35) Dibromofluoromethane	7.519	113	273752	46.60	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.20%
50) Toluene-d8	10.037	98	1051640	47.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.48%
62) 4-Bromofluorobenzene	12.357	95	337364	45.24	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.48%
Target Compounds						
					Qvalue	
16) Acetone	3.767	43	236578	111.17	ug/l	99
25) 2-Butanone	6.765	43	4912	1.48	ug/l #	79
37) Ethyl Acetate	6.854	43	52241	7.25	ug/l #	90
52) Toluene	10.102	92	23929	1.49	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067110.D
Acq On : 17 May 2021 18:35
Operator : JC/MD
Sample : M2387-09DL 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30794DL

Quant Time: May 18 04:11:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

