

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067111.D
 Acq On : 17 May 2021 18:58
 Operator : JC/MD
 Sample : M2387-10 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30795

Quant Time: May 18 04:11:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.600	168	597603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	840542	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	742719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.304	152	337919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	282719	43.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.58%
35) Dibromofluoromethane	7.519	113	258263	46.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.58%
50) Toluene-d8	10.037	98	997957	47.71	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.42%
62) 4-Bromofluorobenzene	12.357	95	335523	47.88	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.76%
Target Compounds						
					Qvalue	
16) Acetone	3.772	43	9264	4.62	ug/l	96
37) Ethyl Acetate	6.862	43	34793	5.14	ug/l	100
52) Toluene	10.102	92	7954	0.53	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067111.D
Acq On : 17 May 2021 18:58
Operator : JC/MD
Sample : M2387-10 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30795

Quant Time: May 18 04:11:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

