

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067114.D
 Acq On : 17 May 2021 20:09
 Operator : JC/MD
 Sample : M2367-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PV-SP2

Quant Time: May 18 04:12:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.599	168	641581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	904114	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.357	117	806678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	333416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	319415	46.08	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.16%
35) Dibromofluoromethane	7.522	113	294200	49.56	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.12%
50) Toluene-d8	10.037	98	1143409	50.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.64%
62) 4-Bromofluorobenzene	12.357	95	365288	48.47	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.94%
Target Compounds						
16) Acetone	3.772	43	30880	14.33	ug/l	99
20) Methylene Chloride	4.488	84	50599	7.24	ug/l	98
43) Isopropyl Acetate	8.104	43	45666	3.78	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067114.D
Acq On : 17 May 2021 20:09
Operator : JC/MD
Sample : M2367-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PV-SP2

Quant Time: May 18 04:12:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

