

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
 Data File : VN067115.D
 Acq On : 17 May 2021 20:33
 Operator : JC/MD
 Sample : M2405-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BUR-1013-1019

Quant Time: May 18 04:12:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.599	168	644776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	915061	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	812336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	351331	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	301812	43.33	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.66%
35) Dibromofluoromethane	7.522	113	276477	46.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.02%
50) Toluene-d8	10.037	98	1079535	47.40	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.80%
62) 4-Bromofluorobenzene	12.357	95	358006	46.93	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.86%
Target Compounds						
					Qvalue	
16) Acetone	3.767	43	1513571	699.13	ug/l	100
20) Methylene Chloride	4.486	84	34277	4.91	ug/l	97
25) 2-Butanone	6.765	43	20363	6.03	ug/l	99
43) Isopropyl Acetate	8.104	43	46894	3.83	ug/l #	92
51) 4-Methyl-2-Pentanone	9.933	43	24435	8.65	ug/l #	86
52) Toluene	10.102	92	316289	19.31	ug/l	98
95) Naphthalene	15.093	128	31259	5.21	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051721\
Data File : VN067115.D
Acq On : 17 May 2021 20:33
Operator : JC/MD
Sample : M2405-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BUR-1013-1019

Quant Time: May 18 04:12:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

