

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077709.D
 Acq On : 17 May 2023 13:21
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 18 01:38:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	364619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	672067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	578245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	224966	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	252265	46.239	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.480%
35) Dibromofluoromethane	8.171	113	212201	48.420	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	10.571	98	814538	48.288	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.853	95	265158	44.508	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077709.D
Acq On : 17 May 2023 13:21
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 18 01:38:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

