

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077715.D
 Acq On : 17 May 2023 15:43
 Operator : JC\MD
 Sample : 02831-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-1

Quant Time: May 18 01:39:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	360050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	639645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	556436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	207273	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	248567	46.140	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.280%
35) Dibromofluoromethane	8.171	113	205063	49.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.320%
50) Toluene-d8	10.571	98	782667	48.750	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.859	95	247992	43.736	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077715.D
Acq On : 17 May 2023 15:43
Operator : JC\MD
Sample : 02831-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-1

Quant Time: May 18 01:39:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

