

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077718.D
 Acq On : 17 May 2023 16:55
 Operator : JC\MD
 Sample : 02831-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-4

Quant Time: May 18 01:40:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	349491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	643127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	564201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	224397	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	248429	47.507	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.020%
35) Dibromofluoromethane	8.177	113	204291	48.713	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.420%
50) Toluene-d8	10.571	98	794244	49.203	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.400%
62) 4-Bromofluorobenzene	12.859	95	259208	45.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077718.D
Acq On : 17 May 2023 16:55
Operator : JC\MD
Sample : 02831-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-4

Quant Time: May 18 01:40:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

