

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077719.D
 Acq On : 17 May 2023 17:19
 Operator : JC\MD
 Sample : 02831-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-5

Quant Time: May 18 01:40:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	350065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	645442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	557039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	205061	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	250788	47.880	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.760%
35) Dibromofluoromethane	8.177	113	207474	49.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.580%
50) Toluene-d8	10.571	98	786533	48.551	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.100%
62) 4-Bromofluorobenzene	12.853	95	244610	42.752	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077719.D
Acq On : 17 May 2023 17:19
Operator : JC\MD
Sample : 02831-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-5

Quant Time: May 18 01:40:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

