

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077721.D
 Acq On : 17 May 2023 18:06
 Operator : JC\MD
 Sample : 02831-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-6

Quant Time: May 18 01:40:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	354086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	651879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	581667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	239012	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	255226	48.174	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.340%
35) Dibromofluoromethane	8.177	113	207352	48.779	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.560%
50) Toluene-d8	10.577	98	809683	49.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.853	95	263805	45.652	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077721.D
Acq On : 17 May 2023 18:06
Operator : JC\MD
Sample : 02831-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-6

Quant Time: May 18 01:40:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

