

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
 Data File : VN077725.D
 Acq On : 17 May 2023 19:42
 Operator : JC\MD
 Sample : 02831-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ME-13D

Quant Time: May 18 01:41:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	347780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	580261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	239621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	249654	47.976	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.960%
35) Dibromofluoromethane	8.171	113	203297	48.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	10.570	98	795312	48.893	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.780%
62) 4-Bromofluorobenzene	12.853	95	263533	45.872	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.740%
Target Compounds						
64) Tetrachloroethene	11.106	164	11054	2.724	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051723\
Data File : VN077725.D
Acq On : 17 May 2023 19:42
Operator : JC\MD
Sample : 02831-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ME-13D

Quant Time: May 18 01:41:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

