

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051820\
 Data File : VN061439.D
 Acq On : 18 May 2020 17:04
 Operator : JC/MD
 Sample : L2679-05MEDL 10X
 Misc : 6.51g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-4-14.0-14.5MEDL

Quant Time: May 19 03:42:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	284275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	507602	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	494249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	228231	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194639	45.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.76%	
35) Dibromofluoromethane	7.55	113	146178	54.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.94%	
50) Toluene-d8	10.06	98	645181	52.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.62%	
62) 4-Bromofluorobenzene	12.37	95	247126	53.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.26%	
Target Compounds						
39) Methylcyclohexane	9.05	83	30384	5.241	ug/l	100
73) Isopropylbenzene	12.22	105	30405	1.659	ug/l	94
78) n-propylbenzene	12.57	91	80067	3.835	ug/l	98
85) sec-Butylbenzene	13.15	105	70139	4.036	ug/l	87
89) n-Butylbenzene	13.59	91	74338	5.303	ug/l #	85
95) Naphthalene	15.10	128	290449	24.138	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051820\
Data File : VN061439.D
Acq On : 18 May 2020 17:04
Operator : JC/MD
Sample : L2679-05MEDL 10X
Misc : 6.51g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-4-14.0-14.5MEDL

Quant Time: May 19 03:42:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

