

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN051820\
 Data File : VN061443.D
 Acq On : 18 May 2020 18:37
 Operator : JC/MD
 Sample : L2663-10 10X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1363

Quant Time: May 19 04:29:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	274219	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	504575	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	499734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	224021	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200299	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
35) Dibromofluoromethane	7.55	113	144509	54.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.34%	
50) Toluene-d8	10.06	98	654774	53.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.82%	
62) 4-Bromofluorobenzene	12.38	95	252470	55.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051820\
Data File : VN061443.D
Acq On : 18 May 2020 18:37
Operator : JC/MD
Sample : L2663-10 10X
Misc : 1.00µL/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1363

Quant Time: May 19 04:29:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

