

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051820\
 Data File : VN061446.D
 Acq On : 18 May 2020 19:47
 Operator : JC/MD
 Sample : PB128923TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128923TB

Quant Time: May 19 04:37:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	256800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	476869	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	484535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	212381	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189253	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.55	113	134272	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
50) Toluene-d8	10.06	98	615049	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
62) 4-Bromofluorobenzene	12.37	95	238645	55.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.26%	
Target Compounds						
16) Acetone	3.77	43	21711	11.214	ug/l	97
20) Methylene Chloride	4.51	84	23585	7.488	ug/l	95
43) Isopropyl Acetate	8.13	43	76086	10.788	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051820\
Data File : VN061446.D
Acq On : 18 May 2020 19:47
Operator : JC/MD
Sample : PB128923TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128923TB

Quant Time: May 19 04:37:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

