

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051820\
 Data File : VN061448.D
 Acq On : 18 May 2020 20:34
 Operator : JC/MD
 Sample : L2665-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AMENDED-COMP-SPLP-LEACHATE

Manual Integrations
 APPROVED

MMDadoda
 5/19/2020 4:24:52 PM

Quant Time: May 19 04:41:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	274502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	508019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	504288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	228401	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	199973	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
35) Dibromofluoromethane	7.54	113	20640	7.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	15.52%	
50) Toluene-d8	10.06	98	652794	52.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.76%	
62) 4-Bromofluorobenzene	12.38	95	251636	54.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.12%	

Target Compounds

					Qvalue
11) Tert butyl alcohol	4.73	59	5663m	11.997	ug/l
16) Acetone	3.78	43	37606	25.611	ug/l
19) Methyl tert-butyl Ether	4.99	73	13218	1.238	ug/l #
20) Methylene Chloride	4.50	84	5156	1.531	ug/l
25) 2-Butanone	6.80	43	12733	6.826	ug/l
39) Methylcyclohexane	9.05	83	20596	3.550	ug/l
67) Ethyl Benzene	11.48	91	33605	1.790	ug/l
68) m/p-Xylenes	11.59	106	12684	1.780	ug/l
69) o-Xylene	11.92	106	13276	1.925	ug/l
73) Isopropylbenzene	12.22	105	19282	1.052	ug/l
80) 1,3,5-Trimethylbenzene	12.71	105	24623	1.640	ug/l
84) 1,2,4-Trimethylbenzene	13.01	105	75622	5.033	ug/l
95) Naphthalene	15.10	128	87980	7.306	ug/l

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051820\
Data File : VN061448.D
Acq On : 18 May 2020 20:34
Operator : JC/MD
Sample : L2665-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AMENDED-COMP-SPLP-LEACHATE

Manual Integrations
APPROVED

MMDadoda
5/19/2020 4:24:52 PM

Quant Time: May 19 04:41:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

