

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051820\
 Data File : VN061469.D
 Acq On : 19 May 2020 8:53
 Operator : JC/MD
 Sample : L2671-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS052MW128-200514

Quant Time: May 19 11:49:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	318951	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	568360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	547909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	245423	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	211112	44.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.70%	
35) Dibromofluoromethane	7.55	113	163279	54.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.68%	
50) Toluene-d8	10.06	98	700518	50.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.44%	
62) 4-Bromofluorobenzene	12.38	95	272279	52.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.54%	

Target Compounds

						Qvalue
27) cis-1,2-Dichloroethene	6.78	96	34166	7.672	ug/l	86
39) Methylcyclohexane	9.05	83	6064	0.934	ug/l #	93
40) Benzene	8.00	78	57928	3.617	ug/l	99
44) Trichloroethene	8.80	130	5431	1.294	ug/l	92
52) Toluene	10.13	92	65941	6.413	ug/l	97
67) Ethyl Benzene	11.48	91	87804	4.305	ug/l	99
68) m/p-Xylenes	11.59	106	102390	13.226	ug/l	99
69) o-Xylene	11.92	106	20778	2.774	ug/l	99
73) Isopropylbenzene	12.22	105	4657	0.236	ug/l	95
78) n-propylbenzene	12.57	91	24054	1.072	ug/l	99
80) 1,3,5-Trimethylbenzene	12.71	105	22205	1.376	ug/l	100
84) 1,2,4-Trimethylbenzene	13.01	105	87967	5.449	ug/l	99
95) Naphthalene	15.10	128	118396	9.150	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051820\
Data File : VN061469.D
Acq On : 19 May 2020 8:53
Operator : JC/MD
Sample : L2671-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS052MW128-200514

Quant Time: May 19 11:49:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

